



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128485

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	11/28/2023	15:40
2	CAL2	1	Water	NA	NA	20.6	7.00	11/28/2023	15:42
3	CAL3	1	Water	NA	NA	20.7	10.02	11/28/2023	15:45
4	ICV	1	Water	NA	NA	20.5	7.01	11/28/2023	15:50
5	CCV1	1	Water	NA	NA	20.6	2.02	11/28/2023	16:10
6	O5561-01	1	Water	NA	NA	20.5	6.91	11/28/2023	16:15
7	O5561-01DUP	1	Water	NA	NA	20.6	6.92	11/28/2023	16:16
8	CCV2	1	Water	NA	NA	20.6	12.02	11/28/2023	16:20

WORKLIST(Hardcopy Internal Chain)

WB 128485

WorkList Name : PH O5561 WorkList ID : 175876 Department : Wet-Chemistry Date : 11-28-2023 15:34:56

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5561-01	A821	Water	pH	Cool 4 deg C	PSEG03	L32	11/28/2023	9040C

Date/Time 11/28/23 15:40
 Raw Sample Received by: JD (CSCM)
 Raw Sample Relinquished by: JD (CSCM)

Date/Time 11/28/23
 Raw Sample Received by: JD (CSCM)
 Raw Sample Relinquished by: JD (CSCM)