


LB128103

SFAM01.1

INSTRUMENT - ID: CV4



Linear

A= 0.0000e+000

B= 2.1965e-004 slope

C= -3.0799e-002 *-1.016e-001*

Rho= 0.9996040

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1/1
0.00	0.000	0.019	0.019	225	0.000	225					11
0.05	0.050					442	0				30
0.20	0.200	0.260	0.060	1324	0.0 %	1324					8
2.50	2.500	2.304	-0.196	10630	0.0 %	10630					2
5.00	5.000	5.124	0.124	23469	0.0 %	23469					1
7.50	7.500	7.555	0.055	34537	0.0 %	34537					1
10.0	10.000	9.938	-0.062	45383	0.0 %	45383					1

LB128503
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	225	-		0 SFAM01.1	PPB	11/30/2023 5:06	Std
	0.2 S01	1324	-		0.2 SFAM01.1	PPB	11/30/2023 5:10	Std
	2.5 S02	10630	-		2.5 SFAM01.1	PPB	11/30/2023 5:12	Std
	5 S03	23469	-		5 SFAM01.1	PPB	11/30/2023 5:14	Std
	7.5 S04	34537	-		7.5 SFAM01.1	PPB	11/30/2023 5:17	Std
	10 S05	45383	-		10 SFAM01.1	PPB	11/30/2023 5:22	Std
ICV095	ICV095	18372	4.0047 -		SFAM01.1	PPB	11/30/2023 5:25	SMPL
ICB095	ICB095	146	0.0013 -		SFAM01.1	PPB	11/30/2023 5:27	SMPL
CCV008	CCV008	23613	5.1559 -		SFAM01.1	PPB	11/30/2023 5:30	SMPL
CCB008	CCB008	170	0.0065 -		SFAM01.1	PPB	11/30/2023 5:32	SMPL
PB157462BL	PBW462	208	0.0149 -		SFAM01.1	PPB	11/30/2023 5:34	SMPL
O5329-01	MA49T3	238	0.0215 -		SFAM01.1	PPB	11/30/2023 5:36	SMPL
O5329-02	MA49T6	248	0.0237 -		SFAM01.1	PPB	11/30/2023 5:39	SMPL
O5329-03	MA49P9	241	0.0221 -		SFAM01.1	PPB	11/30/2023 5:41	SMPL
O5329-04	MA49T5	240	0.0219 -		SFAM01.1	PPB	11/30/2023 5:43	SMPL
O5329-05	MA49W7	472	0.0729 -		SFAM01.1	PPB	11/30/2023 5:45	SMPL
O5329-06	MA49T9	268	0.0281 -		SFAM01.1	PPB	11/30/2023 5:48	SMPL
O5329-07	MA49W2	240	0.0219 -		SFAM01.1	PPB	11/30/2023 5:50	SMPL
O5329-08	MA49W2D	263	0.027 -		SFAM01.1	PPB	11/30/2023 5:52	SMPL
O5329-09	MA49W2S	4560	0.9708 -		SFAM01.1	PPB	11/30/2023 5:55	SMPL
O5329-10	MA49W0	195	0.012 -		SFAM01.1	PPB	11/30/2023 5:57	SMPL
O5329-11	MA49Q2	223	0.0182 -		SFAM01.1	PPB	11/30/2023 5:59	SMPL
O5329-12	MA49Z7	222	0.018 -		SFAM01.1	PPB	11/30/2023 6:01	SMPL
O5329-13	MA49Z8	232	0.0202 -		SFAM01.1	PPB	11/30/2023 6:04	SMPL
O5329-15	MA4A01	241	0.0221 -		SFAM01.1	PPB	11/30/2023 6:06	SMPL
PB157463BL	PBW463	234	0.0206 -		SFAM01.1	PPB	11/30/2023 6:08	SMPL
O5330-01	MA49P8	262	0.0268 -		SFAM01.1	PPB	11/30/2023 6:10	SMPL
O5330-02	MA49T7	266	0.0276 -		SFAM01.1	PPB	11/30/2023 6:13	SMPL
O5330-03	MA49T4	271	0.0287 -		SFAM01.1	PPB	11/30/2023 6:15	SMPL
O5330-04	MA49W6	252	0.0246 -		SFAM01.1	PPB	11/30/2023 6:17	SMPL
CCV009	CCV009	23194	5.0638 -		SFAM01.1	PPB	11/30/2023 6:19	SMPL
CCB009	CCB009	178	0.0083 -		SFAM01.1	PPB	11/30/2023 6:22	SMPL
O5330-05	MA49T8	174	0.0074 -		SFAM01.1	PPB	11/30/2023 6:24	SMPL
O5330-06	MA49W8	212	0.0158 -		SFAM01.1	PPB	11/30/2023 6:26	SMPL
O5330-07	MA49W1	253	0.0248 -		SFAM01.1	PPB	11/30/2023 6:29	SMPL
O5330-08	MA49Q3	212	0.0158 -		SFAM01.1	PPB	11/30/2023 6:31	SMPL
O5330-09	MA49W3	237	0.0213 -		SFAM01.1	PPB	11/30/2023 6:33	SMPL
O5330-10	MA49W3D	233	0.0204 -		SFAM01.1	PPB	11/30/2023 6:35	SMPL
O5330-11	MA49W3S	4826	1.0292 -		SFAM01.1	PPB	11/30/2023 6:38	SMPL
O5330-12	MA4A05	200	0.0131 -		SFAM01.1	PPB	11/30/2023 6:40	SMPL
O5330-13	MA49Z9	238	0.0215 -		SFAM01.1	PPB	11/30/2023 6:42	SMPL
O5330-15	MA4A07	273	0.0292 -		SFAM01.1	PPB	11/30/2023 6:44	SMPL
CCV010	CCV010	22969	5.0144 -		SFAM01.1	PPB	11/30/2023 7:09	SMPL
CCB010	CCB010	293	0.0336 -		SFAM01.1	PPB	11/30/2023 7:12	SMPL
PB157465BL	PBW465	133	-0.0016 -		SFAM01.1	PPB	11/30/2023 7:43	SMPL
O5392-01	MA4A06	333	0.0423 -		SFAM01.1	PPB	11/30/2023 7:46	SMPL

LB128503
INSTRUMENT ID : CV1

O5392-02	MA4A06D	228	0.0193 -	SFAM01.1	PPB	11/30/2023 7:48 SMPL
O5392-03	MA4A06S	4458	0.9484 -	SFAM01.1	PPB	11/30/2023 7:50 SMPL
O5393-01	MA4A00	77	-0.0139 -	SFAM01.1	PPB	11/30/2023 7:52 SMPL
O5393-02	MA4A00D	99	-0.0091 -	SFAM01.1	PPB	11/30/2023 7:55 SMPL
O5393-03	MA4A00S	4480	0.9532 -	SFAM01.1	PPB	11/30/2023 7:57 SMPL
CCV011	CCV011	22892	4.9975 -	SFAM01.1	PPB	11/30/2023 7:59 SMPL
CCB011	CCB011	231	0.0199 -	SFAM01.1	PPB	11/30/2023 8:02 SMPL
PB157466BL	PBW466	195	0.012 -	SFAM01.1	PPB	11/30/2023 8:04 SMPL
O5410-01	COS001	235	0.0208 -	SFAM01.1	PPB	11/30/2023 8:06 SMPL
O5410-02	COS002	315	0.0384 -	SFAM01.1	PPB	11/30/2023 8:08 SMPL
O5410-03	COS003	364	0.0492 -	SFAM01.1	PPB	11/30/2023 8:11 SMPL
O5410-04	COS003D	265	0.0274 -	SFAM01.1	PPB	11/30/2023 8:13 SMPL
O5410-05	COS003S	4599	0.9794 -	SFAM01.1	PPB	11/30/2023 8:15 SMPL
O5410-06	COS004	152	0.0026 -	SFAM01.1	PPB	11/30/2023 8:18 SMPL
O5410-07	COS005	158	0.0039 -	SFAM01.1	PPB	11/30/2023 8:20 SMPL
O5410-08	COS006	237	0.0213 -	SFAM01.1	PPB	11/30/2023 8:22 SMPL
O5410-09	COS007	135	-0.0011 -	SFAM01.1	PPB	11/30/2023 8:24 SMPL
O5410-10	COS008	203	0.0138 -	SFAM01.1	PPB	11/30/2023 8:27 SMPL
O5410-11	COS009	165	0.0054 -	SFAM01.1	PPB	11/30/2023 8:29 SMPL
O5410-12	COS010	164	0.0052 -	SFAM01.1	PPB	11/30/2023 8:31 SMPL
O5410-13	COS011	157	0.0037 -	SFAM01.1	PPB	11/30/2023 8:33 SMPL
O5410-14	COS012	176	0.0079 -	SFAM01.1	PPB	11/30/2023 8:36 SMPL
O5410-15	COS013	200	0.0131 -	SFAM01.1	PPB	11/30/2023 8:38 SMPL
O5410-16	COS014	261	0.0265 -	SFAM01.1	PPB	11/30/2023 8:40 SMPL
O5410-17	COS015	200	0.0131 -	SFAM01.1	PPB	11/30/2023 8:43 SMPL
O5410-18	COS016	201	0.0134 -	SFAM01.1	PPB	11/30/2023 8:45 SMPL
O5410-19	COS017	184	0.0096 -	SFAM01.1	PPB	11/30/2023 8:47 SMPL
CCV012	CCV012	22688	4.9527 -	SFAM01.1	PPB	11/30/2023 8:49 SMPL
CCB012	CCB012	262	0.0268 -	SFAM01.1	PPB	11/30/2023 8:52 SMPL
O5410-20	COS018	123	-0.0038 -	SFAM01.1	PPB	11/30/2023 8:54 SMPL
O5410-21	COS019	150	0.0021 -	SFAM01.1	PPB	11/30/2023 8:56 SMPL
O5410-22	COS020	178	0.0083 -	SFAM01.1	PPB	11/30/2023 8:58 SMPL
CCV013	CCV013	23248	5.0757 -	SFAM01.1	PPB	11/30/2023 9:01 SMPL
CCB013	CCB013	240	0.0219 -	SFAM01.1	PPB	11/30/2023 9:03 SMPL