

LB128510


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

48623



A= 0.0000e+000

B= 1.9534e-004

C= -1.1038e-002

Rho= 0.9961737

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01)
0.0	0.000	-0.007	-0.007	23	0.000	23					
0.2	0.200	0.179	-0.021	974	0.0 %	974					1
2.5	2.500	2.306	-0.194	11862	0.0 %	11862					1.1
5.0	5.000	5.205	0.205	26701	0.0 %	26701					8
7.5	7.500	8.030	0.530	41164	0.0 %	41164					4
10.0	10.000	9.487	-0.513	48623	0.0 %	48623					7

5

LB128510 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Nov 2023 13:05:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	23	-	0.0000	7471B	PPB	30 Nov 2023 13:09:21	S	Std
0.2 - 1	50.2	974	-	0.2000	7471B	PPB	30 Nov 2023 13:11:38	S	Std
2.5 - 1	52.5	11862	-	2.5000	7471B	PPB	30 Nov 2023 13:13:55	S	Std
5.0 - 1	55	26701	-	5.0000	7471B	PPB	30 Nov 2023 13:16:12	S	Std
7.5 - 1	57.5	41164	-	7.5000	7471B	PPB	30 Nov 2023 13:18:31	S	Std
10.0 - 1	510	48623	-	10.0000	7471B	PPB	30 Nov 2023 13:27:55	S	Std
ICV96 - 1	ICV96	21591	4.2065	-	7471B	PPB	30 Nov 2023 13:31:50	U	SMPL
ICB96 - 1	ICB96	-411	-0.0913	-	7471B	PPB	30 Nov 2023 13:34:05	U	SMPL
CCV14 - 1	CCV14	26706	5.2056	-	7471B	PPB	30 Nov 2023 13:36:25	U	SMPL
CCB14 - 1	CCB14	-376	-0.0845	-	7471B	PPB	30 Nov 2023 13:38:40	U	SMPL
CRA - 1	CRA	925	0.1696	-	7471B	PPB	30 Nov 2023 13:41:04	U	SMPL
HighStd - 1	HighStd	49130	9.5859	-	7471B	PPB	30 Nov 2023 13:43:19	U	SMPL
ChkStd - 1	ChkStd	38618	7.5325	-	7471B	PPB	30 Nov 2023 13:45:35	U	SMPL
PB157490BL - 1	PBS	-229	-0.0558	-	7471B	PPB	30 Nov 2023 13:48:03	U	SMPL
PB157490BS - 1	LCSS	20158	3.9266	-	7471B	PPB	30 Nov 2023 13:50:32	U	SMPL
O5547-03 - 1	BCC-8002	-313	-0.0722	-	7471B	PPB	30 Nov 2023 13:52:48	U	SMPL
O5553-01 - 1	HD-01-112723	559	0.0982	-	7471B	PPB	30 Nov 2023 13:55:09	U	SMPL
O5556-01 - 1	SP-1-COMP	16844	3.2792	-	7471B	PPB	30 Nov 2023 13:57:26	U	SMPL
O5560-01 - 1	RT4746	964	0.1773	-	7471B	PPB	30 Nov 2023 13:59:45	U	SMPL
O5562-01 - 1	SU-03-112823	1352	0.2531	-	7471B	PPB	30 Nov 2023 14:02:04	U	SMPL
CCV15 - 1	CCV15	26881	5.2398	-	7471B	PPB	30 Nov 2023 14:04:20	U	SMPL
CCB15 - 1	CCB15	-250	-0.0599	-	7471B	PPB	30 Nov 2023 14:06:35	U	SMPL
O5576-01 - 1	TR-04-112923	1151	0.2138	-	7471B	PPB	30 Nov 2023 14:08:59	U	SMPL
O5577-01 - 1	EO-03-112923	3996	0.7695	-	7471B	PPB	30 Nov 2023 14:15:31	U	SMPL
O5577-01DUP - 1	EO-03-112923DUP	3947	0.7600	-	7471B	PPB	30 Nov 2023 14:17:48	U	SMPL
O5577-01MS - 1	EO-03-112923MS	24971	4.8667	-	7471B	PPB	30 Nov 2023 14:20:05	U	SMPL
O5577-01MSD - 1	EO-03-112923MSD	24987	4.8698	-	7471B	PPB	30 Nov 2023 14:22:22	U	SMPL
O5577-01LX5 - 1		909	0.1665	-	7471B	PPB	30 Nov 2023 14:27:44	U	SMPL
O5577-01A - 1		25473	4.9648	-	7471B	PPB	30 Nov 2023 14:30:00	U	SMPL
CCV16 - 1	CCV16	27534	5.3674	-	7471B	PPB	30 Nov 2023 14:32:17	U	SMPL
CCB16 - 1	CCB16	-216	-0.0532	-	7471B	PPB	30 Nov 2023 14:34:41	U	SMPL