



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128514

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	11/30/2023	12:20
2	O5561-01		1	14	100.00	>212.0	>212.0	11/30/2023	12:50
3	O5571-01		1	12	100.00	>212.0	>212.0	11/30/2023	13:20
4	O5579-04		1	14	100.00	>212.0	>212.0	11/30/2023	13:50
5	O5579-08		1	14	100.00	>212.0	>212.0	11/30/2023	14:20
6	O5579-10		1	12	100.00	>212.0	>212.0	11/30/2023	14:50
7	O5579-12		1	13	100.00	>212.0	>212.0	11/30/2023	15:20
8	O5580-06		1	14	100.00	>212.0	>212.0	11/30/2023	15:50
9	O5580-06DUP		1	14	100.00	>212.0	>212.0	11/30/2023	16:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

6128514

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-11-29

WorkList ID : 175895

Department : Wet-Chemistry

Date : 11-29-2023 08:05:36

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5561-01 Q	A821	Water	Flash Point	Cool 4 deg C	PSEG03	L32	11/28/2023	1010B
O5571-01 C	50138	Water	Flash Point	Cool 4 deg C	PSEG03	L33	11/29/2023	1010B
O5579-04 E	3339	Water	Flash Point	Cool 4 deg C	PSEG03	L33	11/29/2023	1010B
O5579-08 E	3326-TOTE	Water	Flash Point	Cool 4 deg C	PSEG03	L33	11/29/2023	1010B
O5579-10 E	3364-TOTE	Water	Flash Point	Cool 4 deg C	PSEG03	L33	11/29/2023	1010B
O5579-12 E	3363-TOTE	Water	Flash Point	Cool 4 deg C	PSEG03	L33	11/29/2023	1010B
O5580-06 E	223	Water	Flash Point	Cool 4 deg C	PSEG03	L33	11/29/2023	1010B

Date/Time 11/30/2023 12:10
Raw Sample Received by: RA CWC
Raw Sample Relinquished by: RA CWC

Date/Time 11/30/2023 16:30
Raw Sample Received by: RA CWC
Raw Sample Relinquished by: RA CWC