



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB128520

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/01/2023	08:00
2	CAL2	1	Water	NA	NA	20.3	7.01	12/01/2023	08:01
3	CAL3	1	Water	NA	NA	20.4	10.02	12/01/2023	08:05
4	ICV	1	Water	NA	NA	20.4	6.98	12/01/2023	08:10
5	CCV1	1	Water	NA	NA	20.6	2.01	12/01/2023	08:11
6	O5571-01	1	Water	NA	NA	20.4	7.38	12/01/2023	08:15
7	O5571-01DUP	1	Water	NA	NA	20.5	7.39	12/01/2023	08:16
8	O5591-04	1	Water	NA	NA	23.6	6.53	12/01/2023	08:20
9	O5591-05	1	Water	NA	NA	23.2	6.75	12/01/2023	08:24
10	O5591-06	1	Water	NA	NA	23.2	6.74	12/01/2023	08:25
11	O5591-07	1	Water	NA	NA	22.9	6.69	12/01/2023	08:30
12	O5591-08	1	Water	NA	NA	23.1	6.58	12/01/2023	08:35
13	O5591-10	1	Water	NA	NA	22.9	6.53	12/01/2023	08:37
14	O5591-11	1	Water	NA	NA	23.0	5.61	12/01/2023	08:45
15	O5591-12	1	Water	NA	NA	23.0	6.19	12/01/2023	08:50
16	CCV2	1	Water	NA	NA	20.5	12.02	12/01/2023	09:00
17	O5591-13	1	Water	NA	NA	23.1	6.32	12/01/2023	09:10
18	O5591-14	1	Water	NA	NA	23.2	6.47	12/01/2023	09:20
19	O5591-15	1	Water	NA	NA	23.6	7.31	12/01/2023	09:25
20	CCV3	1	Water	NA	NA	20.5	2.01	12/01/2023	09:30

WORKLIST(Hardcopy Internal Chain)

VB 128520

WorkList Name : ph o5591 WorkList ID : 175947 Department : Wet-Chemistry Date : 12-01-2023 07:37:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5571-01	50138	Water	pH	Cool 4 deg C	PSEG03	L33	11/29/2023	9040C
O5591-04	335673	Water	pH	Cool 4 deg C	PSEG03	L33	11/30/2023	9040C
O5591-05	335674	Water	pH	Cool 4 deg C	PSEG03	L33	11/30/2023	9040C
O5591-06	335675	Water	pH	Cool 4 deg C	PSEG03	L33	11/30/2023	9040C
O5591-07	335680	Water	pH	Cool 4 deg C	PSEG03	L33	11/30/2023	9040C
O5591-08	335681	Water	pH	Cool 4 deg C	PSEG03	L33	11/30/2023	9040C
O5591-10	335679	Water	pH	Cool 4 deg C	PSEG03	L33	11/30/2023	9040C
O5591-11	335682	Water	pH	Cool 4 deg C	PSEG03	L33	11/30/2023	9040C
O5591-12	335683	Water	pH	Cool 4 deg C	PSEG03	L33	11/30/2023	9040C
O5591-13	335684	Water	pH	Cool 4 deg C	PSEG03	L33	11/30/2023	9040C
O5591-14	335685	Water	pH	Cool 4 deg C	PSEG03	L33	11/30/2023	9040C
O5591-15	34451	Water	pH	Cool 4 deg C	PSEG03	L33	11/30/2023	9040C

Date/Time 12/01/23 07:40

Raw Sample Received by: JH (aqc)

Raw Sample Relinquished by: JH (aqc)

Date/Time 12/01/23 12:00

Raw Sample Received by: JH (aqc)

Raw Sample Relinquished by: JH (aqc)