



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB128521

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	12/01/2023	09:25
2	CAL2	1	Water	NA	NA	20.6	6.98	12/01/2023	09:26
3	CAL3	1	Water	NA	NA	20.5	10.02	12/01/2023	09:28
4	ICV	1	Water	NA	NA	20.5	7.00	12/01/2023	09:33
5	CCV1	1	Water	NA	NA	20.4	2.01	12/01/2023	09:37
6	O5588-03	1	Solid	20.02	20	24.3	10.50	12/01/2023	09:45
7	O5588-03DUP	1	Solid	20.01	20	24.4	10.51	12/01/2023	09:46
8	O5589-01	1	Solid	20.02	20	23.7	6.42	12/01/2023	09:50
9	O5589-02	1	Solid	20.01	20	23.2	4.58	12/01/2023	10:00
10	CCV2	1	Water	NA	NA	20.5	12.02	12/01/2023	10:10

WORKLIST(Hardcopy Internal Chain)

128627

WorkList Name : CORRSOVITY O5581 WorkList ID : 175968 Department : Wet-Chemistry Date : 12-01-2023 09:09:55

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5588-03	TUDOR	Solid	Corrosivity	Cool 4 deg C	TULL02	L33	11/30/2023	9045D
O5589-01	DRUM-A-OH-OILY-STONE	Solid	Corrosivity	Cool 4 deg C	PSEG03	L33	11/30/2023	9045D
O5589-02	DRUM-B-OH-OIL	Solid	Corrosivity	Cool 4 deg C	PSEG03	L33	11/30/2023	9045D

Date/Time 12/01/23 09:15
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 12/01/23 12:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]