

LB128527


245.1

INSTRUMENT ID: CV4

Linear

 μ Abs.:

41240



A= 0.0000e+000

B= 2.4097e-004

C= -3.4528e-002

Rho= 0.9997300

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.005	-0.005	121	0.000	121					14
0.05	0.050	0.057	0.007	380	0.0 %	380	0				1
0.2	0.200	0.201	0.001	978	0.0 %	978					5
2.5	2.500	2.379	-0.121	10017	0.0 %	10017					3
5.0	5.000	5.157	0.157	21545	0.0 %	21545					1
7.5	7.500	7.558	0.058	31507	0.0 %	31507					1
10.0	10.000	9.903	-0.097	41240	0.0 %	41240					1

LB128527 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 01 Dec 2023 13:22:34

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	121	-	0.0000	245.1	PPB	01 Dec 2023 13:25:35	S	Std
0.05 - 1	50.05	380	-	0.0500	245.1	PPB	01 Dec 2023 13:30:39	S	Std
0.2 - 1	50.2	978	-	0.2000	245.1	PPB	01 Dec 2023 13:32:56	S	Std
2.5 - 1	52.5	10017	-	2.5000	245.1	PPB	01 Dec 2023 13:35:13	S	Std
5.0 - 1	55	21545	-	5.0000	245.1	PPB	01 Dec 2023 13:40:05	S	Std
7.5 - 1	57.5	31507	-	7.5000	245.1	PPB	01 Dec 2023 13:44:55	S	Std
10.0 - 1	511	41240	-	10.0000	245.1	PPB	01 Dec 2023 14:00:30	S	Std
ICV98 - 1	ICV98	16355	3.9065	-245.1	PPB	PPB	01 Dec 2023 14:05:54	U	SMPL
ICB98 - 1	ICB98	117	-0.0063	-245.1	PPB	PPB	01 Dec 2023 14:10:48	U	SMPL
CCV20 - 1	CCV20	22038	5.2760	-245.1	PPB	PPB	01 Dec 2023 14:13:03	U	SMPL
CCB20 - 1	CCB20	26	-0.0283	-245.1	PPB	PPB	01 Dec 2023 14:15:18	U	SMPL
CRA - 1	CRA	314	0.0411	-245.1	PPB	PPB	01 Dec 2023 14:17:36	U	SMPL
HighStd - 1	HighStd	40355	9.6898	-245.1	PPB	PPB	01 Dec 2023 14:19:51	U	SMPL
ChkStd - 1	ChkStd	30402	7.2914	-245.1	PPB	PPB	01 Dec 2023 14:22:06	U	SMPL
PB157516BL - 1	PBW	127	-0.0039	-245.1	PPB	PPB	01 Dec 2023 14:27:04	U	SMPL
PB157516BS - 1	LCSW	16034	3.8292	-245.1	PPB	PPB	01 Dec 2023 14:31:53	U	SMPL
O5480-02 - 1	EFFLUENT-METALS	2099	0.4713	-245.1	PPB	PPB	01 Dec 2023 14:34:09	U	SMPL
O5585-02 - 1	P3874-SP2	249	0.0255	-245.1	PPB	PPB	01 Dec 2023 14:36:27	U	SMPL
O5585-03 - 1	P3874-SP3	372	0.0551	-245.1	PPB	PPB	01 Dec 2023 14:38:46	U	SMPL
O5585-03DUP - 1	P3874-SP3DUP	362	0.0527	-245.1	PPB	PPB	01 Dec 2023 14:41:03	U	SMPL
O5585-03MS - 1	P3874-SP3MS	14697	3.5070	-245.1	PPB	PPB	01 Dec 2023 14:43:18	U	SMPL
CCV21 - 1	CCV21	21558	5.1603	-245.1	PPB	PPB	01 Dec 2023 14:45:33	U	SMPL
CCB21 - 1	CCB21	156	0.0031	-245.1	PPB	PPB	01 Dec 2023 14:50:28	U	SMPL
O5585-03MSD - 1	P3874-SP3MSD	16434	3.9256	-245.1	PPB	PPB	01 Dec 2023 14:52:42	U	SMPL
O5585-03LX5 - 1		-179	-0.0777	-245.1	PPB	PPB	01 Dec 2023 14:54:58	U	SMPL
O5585-03A - 1		17414	4.1617	-245.1	PPB	PPB	01 Dec 2023 14:57:15	U	SMPL
CCV22 - 1	CCV22	21873	5.2362	-245.1	PPB	PPB	01 Dec 2023 14:59:30	U	SMPL
CCB22 - 1	CCB22	186	0.0103	-245.1	PPB	PPB	01 Dec 2023 15:04:22	U	SMPL