


SFAM01.1

LB128534

INSTRUMENT ID: CV7

Linear



A= 0.0000e+000

B= 2.1290e-004 slope

C= -6.2630e-002 y-intercept

Rho= 0.9998612

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/p
0.00	0.000	0.030	0.030	434	0.000	434					=
0.05	0.050					442	0				-8
0.20	0.200	0.184	-0.016	1157	0.0 %	1157					3
2.50	2.500	2.569	0.069	12360	0.0 %	12360					-2
5.00	5.000	4.888	-0.112	23253	0.0 %	23253					6
7.50	7.500	7.471	-0.029	35386	0.0 %	35386					1
10.0	10.000	10.058	0.058	47538	0.0 %	47538					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	434	-		0 SFAM01.1	PPB	12/3/2023 11:19	Std
	0.2 S01	1157	-		0.2 SFAM01.1	PPB	12/3/2023 11:21	Std
	2.5 S02	12360	-		2.5 SFAM01.1	PPB	12/3/2023 11:23	Std
	5 S03	23253	-		5 SFAM01.1	PPB	12/3/2023 11:26	Std
	7.5 S04	35386	-		7.5 SFAM01.1	PPB	12/3/2023 11:28	Std
	10 S05	47538	-		10 SFAM01.1	PPB	12/3/2023 11:30	Std
ICV099	ICV099	18604	3.8982 -		SFAM01.1	PPB	12/3/2023 11:33	SMPL
ICB099	ICB099	-102	-0.0843 -		SFAM01.1	PPB	12/3/2023 11:35	SMPL
CCV023	CCV023	24111	5.0707 -		SFAM01.1	PPB	12/3/2023 11:38	SMPL
CCB023	CCB023	-106	-0.0852 -		SFAM01.1	PPB	12/3/2023 11:40	SMPL
PB157525BL	PBW525	290	-0.0009 -		SFAM01.1	PPB	12/3/2023 11:42	SMPL
O5413-01	COQH0	308	0.0029 -		SFAM01.1	PPB	12/3/2023 11:45	SMPL
O5413-02	COQJ3	713	0.0892 -		SFAM01.1	PPB	12/3/2023 11:52	SMPL
O5413-03	COQJ3D	593	0.0636 -		SFAM01.1	PPB	12/3/2023 11:54	SMPL
O5413-04	COQJ3S	5239	1.0528 -		SFAM01.1	PPB	12/3/2023 11:56	SMPL
O5413-05	COQF0	269	-0.0054 -		SFAM01.1	PPB	12/3/2023 11:58	SMPL
O5413-06	COQG9	243	-0.0109 -		SFAM01.1	PPB	12/3/2023 12:01	SMPL
O5413-07	COQK1	330	0.0076 -		SFAM01.1	PPB	12/3/2023 12:03	SMPL
O5413-08	COQG8	338	0.0093 -		SFAM01.1	PPB	12/3/2023 12:05	SMPL
O5413-09	COQH7	268	-0.0056 -		SFAM01.1	PPB	12/3/2023 12:08	SMPL
O5413-10	COQK0	162	-0.0281 -		SFAM01.1	PPB	12/3/2023 12:10	SMPL
O5413-11	COQH5	144	-0.032 -		SFAM01.1	PPB	12/3/2023 12:12	SMPL
O5413-12	COQJ2	120	-0.0371 -		SFAM01.1	PPB	12/3/2023 12:14	SMPL
O5413-13	MC0QL6	156	-0.0294 -		SFAM01.1	PPB	12/3/2023 12:17	SMPL
O5413-14	COQK2	106	-0.0401 -		SFAM01.1	PPB	12/3/2023 12:19	SMPL
O5413-15	COQF1	110	-0.0392 -		SFAM01.1	PPB	12/3/2023 12:21	SMPL
O5413-16	COQH1	116	-0.0379 -		SFAM01.1	PPB	12/3/2023 12:23	SMPL
O5413-17	COQF7	95	-0.0424 -		SFAM01.1	PPB	12/3/2023 12:26	SMPL
O5413-18	COQF8	132	-0.0345 -		SFAM01.1	PPB	12/3/2023 12:28	SMPL
O5413-19	COQF9	167	-0.0271 -		SFAM01.1	PPB	12/3/2023 12:30	SMPL
CCV024	CCV024	23644	4.9713 -		SFAM01.1	PPB	12/3/2023 12:32	SMPL
CCB024	CCB024	50	-0.052 -		SFAM01.1	PPB	12/3/2023 12:35	SMPL
O5413-20	COQG4	503	0.0445 -		SFAM01.1	PPB	12/3/2023 12:37	SMPL
O5413-21	COQG0	306	0.0025 -		SFAM01.1	PPB	12/3/2023 12:39	SMPL
O5413-22	COQG2	342	0.0102 -		SFAM01.1	PPB	12/3/2023 12:42	SMPL
PB157526BL	PBW526	290	-0.0009 -		SFAM01.1	PPB	12/3/2023 12:44	SMPL
O5427-01	COS021	398	0.0221 -		SFAM01.1	PPB	12/3/2023 12:46	SMPL
O5427-02	COS022	637	0.073 -		SFAM01.1	PPB	12/3/2023 12:48	SMPL
O5427-03	COS022D	573	0.0594 -		SFAM01.1	PPB	12/3/2023 12:51	SMPL
O5427-04	COS022S	5352	1.0768 -		SFAM01.1	PPB	12/3/2023 12:53	SMPL
O5427-05	COS023	624	0.0702 -		SFAM01.1	PPB	12/3/2023 12:55	SMPL
O5427-06	COS024	332	0.0081 -		SFAM01.1	PPB	12/3/2023 12:57	SMPL
O5427-07	COS025	217	-0.0164 -		SFAM01.1	PPB	12/3/2023 13:00	SMPL
O5427-08	COS026	253	-0.0088 -		SFAM01.1	PPB	12/3/2023 13:02	SMPL
O5427-09	COS027	221	-0.0156 -		SFAM01.1	PPB	12/3/2023 13:04	SMPL
O5427-10	COS028	211	-0.0177 -		SFAM01.1	PPB	12/3/2023 13:07	SMPL

O5427-11	COS029	220	-0.0158 -	SFAM01.1	PPB	12/3/2023 13:09 SMPL
O5427-12	COS030	221	-0.0156 -	SFAM01.1	PPB	12/3/2023 13:11 SMPL
O5427-13	COS031	213	-0.0173 -	SFAM01.1	PPB	12/3/2023 13:13 SMPL
O5427-14	COS032	251	-0.0092 -	SFAM01.1	PPB	12/3/2023 13:16 SMPL
O5427-15	COQF1F	209	-0.0181 -	SFAM01.1	PPB	12/3/2023 13:18 SMPL
O5427-16	COQH1F	226	-0.0145 -	SFAM01.1	PPB	12/3/2023 13:20 SMPL
CCV025	CCV025	24141	5.0771 -	SFAM01.1	PPB	12/3/2023 13:22 SMPL
CCB025	CCB025	75	-0.0467 -	SFAM01.1	PPB	12/3/2023 13:25 SMPL
O5427-17	COQF7F	212	-0.0175 -	SFAM01.1	PPB	12/3/2023 13:27 SMPL
O5427-18	COQF8F	222	-0.0154 -	SFAM01.1	PPB	12/3/2023 13:29 SMPL
O5427-19	COQF9F	217	-0.0164 -	SFAM01.1	PPB	12/3/2023 13:31 SMPL
O5427-20	COQG4F	277	-0.0037 -	SFAM01.1	PPB	12/3/2023 13:34 SMPL
O5427-21	COQG0F	169	-0.0266 -	SFAM01.1	PPB	12/3/2023 13:36 SMPL
O5427-22	COQG2F	301	0.0015 -	SFAM01.1	PPB	12/3/2023 13:38 SMPL
PB157527BL	PBW527	16	-0.0592 -	SFAM01.1	PPB	12/3/2023 13:41 SMPL
O5423-08	MC0AK4	143	-0.0322 -	SFAM01.1	PPB	12/3/2023 13:43 SMPL
O5443-01	COQG1	360	0.014 -	SFAM01.1	PPB	12/3/2023 13:45 SMPL
O5443-02	COQH4	-8	-0.0643 -	SFAM01.1	PPB	12/3/2023 13:47 SMPL
O5443-03	COQG5	30	-0.0562 -	SFAM01.1	PPB	12/3/2023 13:50 SMPL
O5443-04	COQG3	48	-0.0524 -	SFAM01.1	PPB	12/3/2023 13:52 SMPL
O5443-05	COQG6	85	-0.0445 -	SFAM01.1	PPB	12/3/2023 13:54 SMPL
O5443-06	COQK5	70	-0.0477 -	SFAM01.1	PPB	12/3/2023 13:56 SMPL
O5443-07	COQK7	297	0.0006 -	SFAM01.1	PPB	12/3/2023 13:59 SMPL
O5443-08	COQH3	15	-0.0594 -	SFAM01.1	PPB	12/3/2023 14:01 SMPL
O5443-09	COQK6	53	-0.0513 -	SFAM01.1	PPB	12/3/2023 14:03 SMPL
O5443-10	MCOQL4	134	-0.0341 -	SFAM01.1	PPB	12/3/2023 14:05 SMPL
O5443-11	COQK8	123	-0.0364 -	SFAM01.1	PPB	12/3/2023 14:08 SMPL
O5443-12	COQL1	114	-0.0384 -	SFAM01.1	PPB	12/3/2023 14:10 SMPL
CCV026	CCV026	23916	5.0292 -	SFAM01.1	PPB	12/3/2023 14:12 SMPL
CCB026	CCB026	90	-0.0435 -	SFAM01.1	PPB	12/3/2023 14:15 SMPL
O5443-13	MCOQL5	116	-0.0379 -	SFAM01.1	PPB	12/3/2023 14:17 SMPL
O5443-14	COQG1D	55	-0.0509 -	SFAM01.1	PPB	12/3/2023 14:19 SMPL
O5443-15	COQG1S	4885	0.9774 -	SFAM01.1	PPB	12/3/2023 14:21 SMPL
O5453-01	MC0AJ0	307	0.0027 -	SFAM01.1	PPB	12/3/2023 14:29 SMPL
O5453-02	MC0AJ0D	284	-0.0022 -	SFAM01.1	PPB	12/3/2023 14:31 SMPL
O5453-03	MC0AJ0S	4712	0.9406 -	SFAM01.1	PPB	12/3/2023 14:33 SMPL
PB157528BL	PBW528	44	-0.0533 -	SFAM01.1	PPB	12/3/2023 14:36 SMPL
O5444-01	COQG1F	183	-0.0237 -	SFAM01.1	PPB	12/3/2023 14:38 SMPL
O5444-02	COQH4F	56	-0.0507 -	SFAM01.1	PPB	12/3/2023 14:40 SMPL
O5444-03	COQG5F	99	-0.0416 -	SFAM01.1	PPB	12/3/2023 14:43 SMPL
O5444-04	COQG3F	106	-0.0401 -	SFAM01.1	PPB	12/3/2023 14:45 SMPL
O5444-05	COQG6F	109	-0.0394 -	SFAM01.1	PPB	12/3/2023 14:47 SMPL
O5444-06	COQK5F	115	-0.0381 -	SFAM01.1	PPB	12/3/2023 14:49 SMPL
O5444-07	COQK7F	208	-0.0183 -	SFAM01.1	PPB	12/3/2023 14:52 SMPL
O5444-08	COQH3F	24	-0.0575 -	SFAM01.1	PPB	12/3/2023 14:54 SMPL
O5444-09	COQK6F	45	-0.053 -	SFAM01.1	PPB	12/3/2023 14:56 SMPL
O5444-10	COQL4F	75	-0.0467 -	SFAM01.1	PPB	12/3/2023 14:58 SMPL

LB128534

INSTRUMENT ID : CV1

O5444-11	COQK8F	67	-0.0484 -	SFAM01.1	PPB	12/3/2023 15:01 SMPL
O5444-12	COQL1F	127	-0.0356 -	SFAM01.1	PPB	12/3/2023 15:03 SMPL
O5444-13	MC0QL5F	69	-0.0479 -	SFAM01.1	PPB	12/3/2023 15:05 SMPL
CCV027	CCV027	24024	5.0522 -	SFAM01.1	PPB	12/3/2023 15:08 SMPL
CCB027	CCB027	124	-0.0362 -	SFAM01.1	PPB	12/3/2023 15:10 SMPL
O5444-14	COQG1FD	38	-0.0545 -	SFAM01.1	PPB	12/3/2023 15:12 SMPL
O5444-15	COQG1FS	4719	0.9421 -	SFAM01.1	PPB	12/3/2023 15:14 SMPL
O5454-01	MC0AH9	629	0.0713 -	SFAM01.1	PPB	12/3/2023 15:17 SMPL
O5454-02	MC0AH9D	266	-0.006 -	SFAM01.1	PPB	12/3/2023 15:19 SMPL
O5454-03	MC0AH9S	4830	0.9657 -	SFAM01.1	PPB	12/3/2023 15:21 SMPL
O5457-08	XA049	19047	3.9926 -	SFAM01.1	PPB	12/3/2023 15:28 SMPL
CCV028	CCV028	24933	5.2457 -	SFAM01.1	PPB	12/3/2023 15:30 SMPL
CCB028	CCB028	36	-0.055 -	SFAM01.1	PPB	12/3/2023 15:33 SMPL