

LB128550


SFAM01.1

INSTRUMENT ID: (V1

Linear

 μ Abs.:

47273



A= 0.0000e+000

B= 2.1311e-004 slope

C= -1.2186e-001 $r^2 = 0.9999$

Rho= 0.9999028

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.00	0.000	-0.054	-0.054	319	0.000	319					
0.05	0.050					442	0				
0.20	0.200	0.219	0.019	1598	0.0 %	1598					
2.50	2.500	2.555	0.055	12563	0.0 %	12563					
5.00	5.000	4.955	-0.045	23823	0.0 %	23823					
7.50	7.500	7.573	0.073	36106	0.0 %	36106					
10.0	10.000	9.952	-0.048	47273	0.0 %	47273					

901
= 10
2
1
1
0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0	319	-		0 SFAM01.1	PPB	12/5/2023 7:52	Std
	0.2	1598	-		0.2 SFAM01.1	PPB	12/5/2023 7:57	Std
	2.5	12563	-		2.5 SFAM01.1	PPB	12/5/2023 7:59	Std
	5	23823	-		5 SFAM01.1	PPB	12/5/2023 8:01	Std
	7.5	36106	-		7.5 SFAM01.1	PPB	12/5/2023 8:06	Std
	10	47273	-		10 SFAM01.1	PPB	12/5/2023 8:11	Std
ICV003	ICV003	20386	4.2225 -		SFAM01.1	PPB	12/5/2023 8:14	SMPL
ICB003	ICB003	251	-0.0684 -		SFAM01.1	PPB	12/5/2023 8:19	SMPL
CCV039	CCV039	24714	5.1448 -		SFAM01.1	PPB	12/5/2023 8:22	SMPL
CCB039	CCB039	432	-0.0298 -		SFAM01.1	PPB	12/5/2023 8:24	SMPL
PB157554BL	PBS554	309	-0.056 -		SFAM01.1	PPB	12/5/2023 8:26	SMPL
O5423-01	MCOAB2	3800	0.6879 -		SFAM01.1	PPB	12/5/2023 8:28	SMPL
O5423-02	MCOAB3	1639	0.2274 -		SFAM01.1	PPB	12/5/2023 8:31	SMPL
O5423-03	MCOAB4	1070	0.1062 -		SFAM01.1	PPB	12/5/2023 8:33	SMPL
O5423-04	MCOAC0	686	0.0243 -		SFAM01.1	PPB	12/5/2023 8:35	SMPL
O5423-05	MCOAC1	773	0.0429 -		SFAM01.1	PPB	12/5/2023 8:37	SMPL
O5423-06	MCOAK2	1111	0.1149 -		SFAM01.1	PPB	12/5/2023 8:40	SMPL
O5423-07	MCOAK3	1139	0.1209 -		SFAM01.1	PPB	12/5/2023 8:42	SMPL
O5423-09	MCOAG6	1295	0.1541 -		SFAM01.1	PPB	12/5/2023 8:44	SMPL
O5423-10	MCOAG7	1541	0.2065 -		SFAM01.1	PPB	12/5/2023 8:47	SMPL
O5423-11	MCOAH0	695	0.0262 -		SFAM01.1	PPB	12/5/2023 8:49	SMPL
O5423-12	MCOAH2	908	0.0716 -		SFAM01.1	PPB	12/5/2023 8:51	SMPL
O5423-13	MCOAH2D	911	0.0723 -		SFAM01.1	PPB	12/5/2023 8:53	SMPL
O5423-14	MCOAH2S	14920	3.0577 -		SFAM01.1	PPB	12/5/2023 8:58	SMPL
O5423-15	MCOAJ9	521	-0.0108 -		SFAM01.1	PPB	12/5/2023 9:01	SMPL
O5423-16	MCOAH4	486	-0.0183 -		SFAM01.1	PPB	12/5/2023 9:03	SMPL
O5423-17	MCOAH5	466	-0.0226 -		SFAM01.1	PPB	12/5/2023 9:05	SMPL
O5423-18	MCOAH6	515	-0.0121 -		SFAM01.1	PPB	12/5/2023 9:07	SMPL
O5423-19	MCOAH7	539	-0.007 -		SFAM01.1	PPB	12/5/2023 9:10	SMPL
O5423-20	MCOAH8	571	-0.0002 -		SFAM01.1	PPB	12/5/2023 9:12	SMPL
CCV040	CCV040	24475	5.0939 -		SFAM01.1	PPB	12/5/2023 9:14	SMPL
CCB040	CCB040	212	-0.0767 -		SFAM01.1	PPB	12/5/2023 9:16	SMPL
O5423-21	MCOAH9	644	0.0154 -		SFAM01.1	PPB	12/5/2023 9:19	SMPL
O5423-22	MCOAJ0	953	0.0812 -		SFAM01.1	PPB	12/5/2023 9:21	SMPL
PB157555BL	PBS555	205	-0.0782 -		SFAM01.1	PPB	12/5/2023 9:23	SMPL
O5475-01	MCOAA2	587	0.0032 -		SFAM01.1	PPB	12/5/2023 9:26	SMPL
O5475-02	MCOAA3	609	0.0079 -		SFAM01.1	PPB	12/5/2023 9:28	SMPL
O5475-03	MCOAA4	598	0.0056 -		SFAM01.1	PPB	12/5/2023 9:30	SMPL
O5475-04	MCOAJ1	1418	0.1803 -		SFAM01.1	PPB	12/5/2023 9:32	SMPL
O5475-05	MCOAJ2	766	0.0414 -		SFAM01.1	PPB	12/5/2023 9:35	SMPL
O5475-06	MCOAJ3	1452	0.1876 -		SFAM01.1	PPB	12/5/2023 9:37	SMPL
O5475-07	MCOAJ4	959	0.0825 -		SFAM01.1	PPB	12/5/2023 9:39	SMPL
O5475-08	MCOAJ6	2923	0.501 -		SFAM01.1	PPB	12/5/2023 9:41	SMPL
O5475-09	MCOAJ7	1236	0.1415 -		SFAM01.1	PPB	12/5/2023 9:44	SMPL
O5475-10	MCOAJ8	490	-0.0174 -		SFAM01.1	PPB	12/5/2023 9:46	SMPL
O5475-11	MCOAK0	906	0.0712 -		SFAM01.1	PPB	12/5/2023 9:48	SMPL

INSTRUMENT ID : CV1

O5475-12	MC0AK1	609	0.0079 -	SFAM01.1	PPB	12/5/2023 9:50 SMPL
O5475-13	MC0AA0	3427	0.6085 -	SFAM01.1	PPB	12/5/2023 9:53 SMPL
O5475-14	MC0AA5	5359	1.0202 -	SFAM01.1	PPB	12/5/2023 9:55 SMPL
O5475-15	MC0AA6	6219	1.2034 -	SFAM01.1	PPB	12/5/2023 9:57 SMPL
O5475-16	MC0AA7	4624	0.8635 -	SFAM01.1	PPB	12/5/2023 10:00 SMPL
O5475-17	MC0AA8	2050	0.315 -	SFAM01.1	PPB	12/5/2023 10:02 SMPL
CCV041	CCV041	25252	5.2595 -	SFAM01.1	PPB	12/5/2023 10:04 SMPL
CCB041	CCB041	474	-0.0208 -	SFAM01.1	PPB	12/5/2023 10:06 SMPL
O5475-18	MC0AA9	2788	0.4723 -	SFAM01.1	PPB	12/5/2023 10:09 SMPL
O5475-19	MC0AA9D	3023	0.5224 -	SFAM01.1	PPB	12/5/2023 10:11 SMPL
O5475-20	MC0AA9S	14744	3.0202 -	SFAM01.1	PPB	12/5/2023 10:13 SMPL
O5475-21	MC0AB0	29735	6.2148 -	SFAM01.1	PPB	12/5/2023 10:15 SMPL
O5475-22	MC0AB1	19160	3.9613 -	SFAM01.1	PPB	12/5/2023 10:18 SMPL
PB157556BL	PBS556	181	-0.0833 -	SFAM01.1	PPB	12/5/2023 10:20 SMPL
O5457-04	XA045	264072	HIGH -	SFAM01.1	PPB	12/5/2023 10:22
O5493-01	MEZZC1	69027	14.5882 -	SFAM01.1	PPB	12/5/2023 10:24 SMPL
O5493-02	MEZZC2	81743	17.2981 -	SFAM01.1	PPB	12/5/2023 10:27 SMPL
O5493-03	MEZZC3	68935	14.5686 -	SFAM01.1	PPB	12/5/2023 10:29 SMPL
O5493-04	MEZZC3D	61394	12.9616 -	SFAM01.1	PPB	12/5/2023 10:31 SMPL
O5493-05	MEZZC3S	82853	17.5346 -	SFAM01.1	PPB	12/5/2023 10:34 SMPL
O5498-01	MC0AB6	3462	0.6159 -	SFAM01.1	PPB	12/5/2023 10:36 SMPL
O5498-02	MC0AB7	4956	0.9343 -	SFAM01.1	PPB	12/5/2023 10:38 SMPL
O5498-03	MC0AC2	2986	0.5145 -	SFAM01.1	PPB	12/5/2023 10:41 SMPL
O5498-04	MC0AC2D	2906	0.4974 -	SFAM01.1	PPB	12/5/2023 10:43 SMPL
O5498-05	MC0AC2S	14821	3.0366 -	SFAM01.1	PPB	12/5/2023 10:46 SMPL
CCV042	CCV042	24223	5.0402 -	SFAM01.1	PPB	12/5/2023 10:53 SMPL
CCB042	CCB042	369	-0.0432 -	SFAM01.1	PPB	12/5/2023 10:56 SMPL
O5457-04DLX10	XA045	38518	8.0866 -	SFAM01.1	PPB	12/5/2023 11:00 SMPL
O5493-03DLX5	MEZZC3	14582	2.9857 -	SFAM01.1	PPB	12/5/2023 11:16 SMPL
O5493-04DLX5	MEZZC3D	12219	2.4821 -	SFAM01.1	PPB	12/5/2023 11:18 SMPL
O5493-05DLX5	MEZZC3S	19914	4.1219 -	SFAM01.1	PPB	12/5/2023 11:26 SMPL
O5493-01DLX5	MEZZC1	14450	2.9575 -	SFAM01.1	PPB	12/5/2023 11:29 SMPL
O5493-02DLX5	MEZZC2	17030	3.5073 -	SFAM01.1	PPB	12/5/2023 11:31 SMPL
O5457-04DLX25	XA045	14805	3.0332 -	SFAM01.1	PPB	12/5/2023 11:40 SMPL
CCV043	CCV043	24756	5.1538 -	SFAM01.1	PPB	12/5/2023 11:42 SMPL
CCB043	CCB043	317	-0.0543 -	SFAM01.1	PPB	12/5/2023 11:45 SMPL