



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 12/04/2023 17:15 TEMP1 OUT: 103 °C 12/05/2023 07:30
TEMP2 IN: 104 °C 12/05/2023 08:00 TEMP2 OUT: 103 °C 12/05/2023 09:30
TEMP3 IN: 540 °C 12/05/2023 10:00 TEMP3 OUT: 540 °C 12/05/2023 11:30
TEMP4 IN: 560 °C 12/05/2023 12:00 TEMP4 OUT: 560 °C 12/05/2023 13:30

Run Number: LB128552

SUPERVISOR: Iwona

ANALYST: JIGNESH

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB128552BL	89.6371	89.6371	100	89.6371	89.6371	89.6371	89.6371	0.0000	0
2	O5620-01	89.9252	89.9252	100	91.0557	91.0557	90.4490	90.4490	0.6067	6067
3	O5620-01DUP	89.8822	89.8822	100	91.0121	91.0121	90.4104	90.4104	0.6017	6017

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105 °C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105 °C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

WORKLIST(Hardcopy Internal Chain)

AD 128552

WorkList Name : TVS O5620 WATER WorkList ID : 176026 Department : Wet-Chemistry Date : 12-04-2023 16:30:34

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5620-01	P OILY-WATER-TOTE	Water	TVS	Cool 4 deg C	PSEG03	L41	12/04/2023	SM2540 G

Date/Time 12/04/23 16:40

Raw Sample Received by: JDCSMD

Raw Sample Relinquished by: JDCSMD

Date/Time 12/04/23 18:00

Raw Sample Received by: JDCSMD

Raw Sample Relinquished by: JDCSMD