



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB128553

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	12/05/2023	08:05
2	CAL2	1	Water	NA	NA	20.6	7.00	12/05/2023	08:06
3	CAL3	1	Water	NA	NA	20.5	10.02	12/05/2023	08:07
4	ICV	1	Water	NA	NA	20.5	7.01	12/05/2023	08:10
5	CCV1	1	Water	NA	NA	20.6	2.01	12/05/2023	08:15
6	O5620-01	1	Water	NA	NA	20.2	6.39	12/05/2023	08:20
7	O5620-01DUP	1	Water	NA	NA	20.3	6.40	12/05/2023	08:21
8	CCV2	1	Water	NA	NA	20.5	12.02	12/05/2023	08:25

VP 128553

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 05620 WorkList ID : 176028 Department : Wet-Chemistry Date : 12-05-2023 07:42:01

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5620-01	P OILY-WATER-TOTE	Water	pH	Cool 4 deg C	PSEG03	L41	12/04/2023	9040C

Date/Time 12/05/23 08:00
Raw Sample Received by: JDC (WC)
Raw Sample Relinquished by: JDC (WC)

Date/Time 12/05/23 09:40
Raw Sample Received by: JDC (WC)
Raw Sample Relinquished by: JDC (WC)