

LB128564


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

56625



A= 0.0000e+000

B= 1.7812e-004

C= 5.5954e-002

Rho= 0.9996766

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.069	0.069	74	0.000	74	0				1
0.2	0.200	0.240	0.040	1033	0.0 %	1033					20
2.5	2.500	2.447	-0.053	13426	0.0 %	13426					2
5.0	5.000	4.944	-0.056	27444	0.0 %	27444					1
7.5	7.500	7.357	-0.143	40990	0.0 %	40990					2
10.0	10.000	10.142	0.142	56625	0.0 %	56625					1

LB128564 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Dec 2023 12:40:06

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	74	-	0.000	7470A	PPB	05 Dec 2023 13:10:30	S	Std
0.2 - 1	51.2	1033	-	0.200	7470A	PPB	05 Dec 2023 13:12:47	S	Std
2.5 - 1	62.5	13426	-	2.500	7470A	PPB	05 Dec 2023 13:15:04	S	Std
5.0 - 1	55	27444	-	5.000	7470A	PPB	05 Dec 2023 13:17:21	S	Std
7.5 - 1	59.5	40990	-	7.500	7470A	PPB	05 Dec 2023 13:19:39	S	Std
10.0 - 1	116	56625	-	10.000	7470A	PPB	05 Dec 2023 13:21:57	S	Std
ICV04 - 1	ICV04	21753	3.9306	-	7470A	PPB	05 Dec 2023 13:24:53	U	SMPL
ICB04 - 1	ICB04	-306	0.0014	-	7470A	PPB	05 Dec 2023 13:27:09	U	SMPL
CCV44 - 1	CCV44	26080	4.7014	-	7470A	PPB	05 Dec 2023 13:45:03	U	SMPL
CCB44 - 1	CCB44	-326	-0.0021	-	7470A	PPB	05 Dec 2023 13:47:17	U	SMPL
CRA - 1	CRA	915	0.2189	-	7470A	PPB	05 Dec 2023 13:49:35	U	SMPL
HighStd - 1	HighStd	56627	10.1424	-	7470A	PPB	05 Dec 2023 13:52:55	U	SMPL
ChkStd - 1	ChkStd	39623	7.1136	-	7470A	PPB	05 Dec 2023 13:55:10	U	SMPL
PB157510BL - 1	PBW	-479	-0.0294	-	7470A	PPB	05 Dec 2023 13:57:28	U	SMPL
PB157510BS - 1	LCSW	22570	4.0761	-	7470A	PPB	05 Dec 2023 14:02:04	U	SMPL
O5563-01 - 1	M-111723-103-01C	-356	-0.0075	-	7470A	PPB	05 Dec 2023 14:04:20	U	SMPL
O5563-02 - 1	M-111723-103-03B	-54	0.0463	-	7470A	PPB	05 Dec 2023 14:06:39	U	SMPL
O5563-03 - 1	M-111723-103-04	-15	0.0533	-	7470A	PPB	05 Dec 2023 14:08:56	U	SMPL
O5573-02 - 1	BLUE-VARIOUS-OIL-DRUM	607	0.1641	-	7470A	PPB	05 Dec 2023 14:11:13	U	SMPL
O5573-04 - 1	2943	1790	0.3748	-	7470A	PPB	05 Dec 2023 14:13:28	U	SMPL
CCV45 - 1	CCV45	29399	5.2925	-	7470A	PPB	05 Dec 2023 14:18:27	U	SMPL
CCB45 - 1	CCB45	-224	0.0161	-	7470A	PPB	05 Dec 2023 14:20:41	U	SMPL
O5573-06 - 1	2945	253	0.1010	-	7470A	PPB	05 Dec 2023 14:22:59	U	SMPL
O5573-08 - 1	2946	278	0.1055	-	7470A	PPB	05 Dec 2023 14:25:14	U	SMPL
O5573-10 - 1	2947	414	0.1297	-	7470A	PPB	05 Dec 2023 14:27:29	U	SMPL
O5574-02 - 1	P001-TS01-02	167	0.0857	-	7470A	PPB	05 Dec 2023 14:29:44	U	SMPL
O5574-03 - 1	P001-WC04-01	309	0.1110	-	7470A	PPB	05 Dec 2023 14:32:00	U	SMPL
O5574-06 - 1	P001-TS01-01	170	0.0862	-	7470A	PPB	05 Dec 2023 14:34:17	U	SMPL
O5574-06DUP - 1	P001-TS01-01DUP	136	0.0802	-	7470A	PPB	05 Dec 2023 14:36:33	U	SMPL
O5574-07 - 1	P001-WC03-01	246	0.0998	-	7470A	PPB	05 Dec 2023 14:38:50	U	SMPL
O5574-11 - 1	O5574-06MS	18394	3.3323	-	7470A	PPB	05 Dec 2023 14:41:07	U	SMPL
O5574-12 - 1	O5574-06MSD	17870	3.2390	-	7470A	PPB	05 Dec 2023 14:46:23	U	SMPL
CCV46 - 1	CCV46	25186	4.5421	-	7470A	PPB	05 Dec 2023 15:11:20	U	SMPL
CCB46 - 1	CCB46	-405	-0.0162	-	7470A	PPB	05 Dec 2023 15:13:36	U	SMPL
O5579-02 - 1	3344	17	0.0590	-	7470A	PPB	05 Dec 2023 15:15:54	U	SMPL
O5579-06 - 1	3330	1881	0.3910	-	7470A	PPB	05 Dec 2023 15:18:10	U	SMPL
O5580-02 - 1	218	-53	0.0465	-	7470A	PPB	05 Dec 2023 15:20:27	U	SMPL
O5580-04 - 1	222	188	0.0894	-	7470A	PPB	05 Dec 2023 15:22:44	U	SMPL
PB157511BL - 1	PBW	216	0.0944	-	7470A	PPB	05 Dec 2023 15:27:45	U	SMPL
PB157511BS - 1	LCSW	18742	3.3943	-	7470A	PPB	05 Dec 2023 15:40:29	U	SMPL
O5579-04 - 1	3339	61	0.0668	-	7470A	PPB	05 Dec 2023 15:42:45	U	SMPL
O5579-04DUP - 1	3339DUP	237	0.0982	-	7470A	PPB	05 Dec 2023 15:45:05	U	SMPL
O5579-04MS - 1	3339MS	14146	2.5757	-	7470A	PPB	05 Dec 2023 15:50:08	U	SMPL
O5579-04MSD - 1	3339MSD	14605	2.6574	-	7470A	PPB	05 Dec 2023 15:52:24	U	SMPL
CCV47 - 1	CCV47	30302	5.4534	-	7470A	PPB	05 Dec 2023 16:01:02	U	SMPL
CCB47 - 1	CCB47	-458	-0.0256	-	7470A	PPB	05 Dec 2023 16:03:17	U	SMPL
O5579-08 - 1	3326-TOTE	29	0.0611	-	7470A	PPB	05 Dec 2023 16:05:34	U	SMPL
O5579-10 - 1	3364-TOTE	-134	0.0321	-	7470A	PPB	05 Dec 2023 16:07:50	U	SMPL
O5579-12 - 1	3363-TOTE	-12	0.0538	-	7470A	PPB	05 Dec 2023 16:10:07	U	SMPL
O5580-06 - 1	223	106	0.0748	-	7470A	PPB	05 Dec 2023 16:12:22	U	SMPL
PB157512BL - 1	PBW	78	0.0698	-	7470A	PPB	05 Dec 2023 16:14:38	U	SMPL
PB157512BS - 1	LCSW	21984	3.9718	-	7470A	PPB	05 Dec 2023 16:22:42	U	SMPL
O5592-01 - 1	MW04S	-389	-0.0133	-	7470A	PPB	05 Dec 2023 16:24:58	U	SMPL
O5594-01 - 1	928	561	0.1559	-	7470A	PPB	05 Dec 2023 16:32:16	U	SMPL
O5594-01DUP - 1	928DUP	473	0.1402	-	7470A	PPB	05 Dec 2023 16:34:33	U	SMPL
O5594-01MS - 1	928MS	20197	3.6535	-	7470A	PPB	05 Dec 2023 16:39:37	U	SMPL
CCV48 - 1	CCV48	29044	5.2293	-	7470A	PPB	05 Dec 2023 16:44:30	U	SMPL
CCB48 - 1	CCB48	-405	-0.0162	-	7470A	PPB	05 Dec 2023 16:46:46	U	SMPL
O5594-01MSD - 1	928MSD	21449	3.8765	-	7470A	PPB	05 Dec 2023 16:49:04	U	SMPL
O5594-02 - 1	929	861	0.2093	-	7470A	PPB	05 Dec 2023 16:51:20	U	SMPL
O5594-03 - 1	930	575	0.1584	-	7470A	PPB	05 Dec 2023 16:53:38	U	SMPL
O5594-04 - 1	931	349	0.1181	-	7470A	PPB	05 Dec 2023 16:55:54	U	SMPL
O5406-10 - 1		99	0.0736	-	7470A	PPB	05 Dec 2023 16:58:09	U	SMPL
CCV49 - 1	CCV49	29019	5.2248	-	7470A	PPB	05 Dec 2023 17:03:22	U	SMPL
CCB49 - 1	CCB49	-327	-0.0023	-	7470A	PPB	05 Dec 2023 17:05:39	U	SMPL
PB157480TB - 1		-55	0.0462	-	7470A	PPB	05 Dec 2023 17:07:59	U	SMPL
PB157481TB - 1		-54	0.0463	-	7470A	PPB	05 Dec 2023 17:10:18	U	SMPL
O5574-06LX5 - 1		92	0.0723	-	7470A	PPB	05 Dec 2023 17:12:36	U	SMPL
O5574-06A - 1		19192	3.4745	-	7470A	PPB	05 Dec 2023 17:17:48	U	SMPL
O5579-04LX5 - 1		-132	0.0324	-	7470A	PPB	05 Dec 2023 17:20:04	U	SMPL
O5579-04A - 1		19050	3.4492	-	7470A	PPB	05 Dec 2023 17:22:22	U	SMPL
O5594-01LX5 - 1		-194	0.0214	-	7470A	PPB	05 Dec 2023 17:24:39	U	SMPL
O5594-01A - 1		17963	3.2555	-	7470A	PPB	05 Dec 2023 17:26:57	U	SMPL
CCV50 - 1	CCV50	26497	4.7756	-	7470A	PPB	05 Dec 2023 17:34:25	U	SMPL
CCB50 - 1	CCB50	-404	-0.0160	-	7470A	PPB	05 Dec 2023 17:36:41	U	SMPL