



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB128571

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	12/06/2023	08:45
2	CAL2	1	Water	NA	NA	20.5	7.00	12/06/2023	08:46
3	CAL3	1	Water	NA	NA	20.5	10.02	12/06/2023	08:50
4	ICV	1	Water	NA	NA	20.5	7.00	12/06/2023	10:00
5	CCV1	1	Water	NA	NA	20.6	2.01	12/06/2023	10:05
6	O5635-08	1	Water	NA	NA	20.7	6.44	12/06/2023	10:15
7	O5635-08DUP	1	Water	NA	NA	20.8	6.45	12/06/2023	10:16
8	CCV2	1	Water	NA	NA	20.6	12.02	12/06/2023	10:20

WORKLIST(Hardcopy Internal Chain)

VB128571

WorkList Name : corrosivity O5635 WATER WorkList ID : 176076 Department : Wet-Chemistry Date : 12-06-2023 08:20:36

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5635-08	COMP-3 A	Water	Corrosivity	Cool 4 deg C	PSEG03	L31	12/05/2023	9040C

Date/Time 12/06/23 08:30
 Raw Sample Received by: TD (OC)
 Raw Sample Relinquished by: Cl Sm

Date/Time 12/06/23 11:00
 Raw Sample Received by: Cl Sm
 Raw Sample Relinquished by: TD (OC)