



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128576

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	12/06/2023	09:10
2	O5591-03		1	14	100.00	>212.0	>212.0	12/06/2023	09:40
3	O5591-03DUP		1	14	100.00	>212.0	>212.0	12/06/2023	10:10
4	O5591-09		1	12	100.00	>212.0	>212.0	12/06/2023	10:40
5	O5591-15		1	13	100.00	>212.0	>212.0	12/06/2023	11:10
6	O5603-01		1	14	100.00	>212.0	>212.0	12/06/2023	12:45
7	O5616-01		1	12	100.00	>212.0	>212.0	12/06/2023	13:15
8	O5616-02		1	12	100.00	>212.0	>212.0	12/06/2023	13:45
9	O5616-03		1	14	100.00	>212.0	>212.0	12/06/2023	14:15
10	O5620-01		1	14	100.00	>212.0	>212.0	12/06/2023	14:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16128576

WorkList Name : fp-12-05

WorkList ID : 176039

Department : Wet-Chemistry

Date : 12-05-2023 08:33:12

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5591-03 B	335673-74-75-80-81	Water	Flash Point	Cool 4 deg C	PSEG03	L33	11/30/2023	1010B
O5591-09 B	335679-82-83-84-85	Water	Flash Point	Cool 4 deg C	PSEG03	L33	11/30/2023	1010B
O5591-15 B	34451	Water	Flash Point	Cool 4 deg C	PSEG03	L33	11/30/2023	1010B
O5603-01 C	34400	Water	Flash Point	Cool 4 deg C	PSEG03	L31	12/01/2023	1010B
O5616-01 D	1148	Water	Flash Point	Cool 4 deg C	PSEG03	L41	12/04/2023	1010B
O5616-02 B	1149	Water	Flash Point	Cool 4 deg C	PSEG03	L41	12/04/2023	1010B
O5616-03 D	1151	Water	Flash Point	Cool 4 deg C	PSEG03	L41	12/04/2023	1010B
O5620-01 S	OILY-WATER-TOTE	Water	Flash Point	Cool 4 deg C	PSEG03	L41	12/04/2023	1010B

Date/Time 12/06/2023 08:30
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: 26102C1

Date/Time 12/06/2023 14:50
 Raw Sample Received by: 701 CWC
 Raw Sample Relinquished by: RM CWC