


LB128528

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01)
0.0	0.000	-0.077	-0.077	35	0.000	35	0				
0.2	0.200	0.173	-0.027	1144	0.0 %	1144					-14
2.5	2.500	2.525	0.025	11604	0.0 %	11604					1
5.0	5.000	5.041	0.041	22794	0.0 %	22794					1
7.5	7.500	7.848	0.348	35276	0.0 %	35276					5
10.0	10.000	9.690	-0.310	43464	0.0 %	43464					-3

LB128578 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Dec 2023 12:14:45

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	35	-	0.00007470A	PPB	06 Dec 2023 12:20:58	S	Std
0.2 - 1	50.2	1144	-	0.20007470A	PPB	06 Dec 2023 12:23:15	S	Std
2.5 - 1	52.5	11604	-	2.50007470A	PPB	06 Dec 2023 12:25:32	S	Std
5.0 - 1	53	22794	-	5.00007470A	PPB	06 Dec 2023 12:27:49	S	Std
7.5 - 1	57.5	35276	-	7.50007470A	PPB	06 Dec 2023 12:35:33	S	Std
10.0 - 1	516	43464	-	10.00007470A	PPB	06 Dec 2023 12:43:02	S	Std
ICV05 - 1	ICV05	17968	3.9560	-7470A	PPB	06 Dec 2023 12:48:40	U	SMPL
ICB05 - 1	ICB05	-33	-0.0921	-7470A	PPB	06 Dec 2023 12:50:54	U	SMPL
CCV50 - 1	CCV50	21679	4.7906	-7470A	PPB	06 Dec 2023 12:53:12	U	SMPL
CCB50 - 1	CCB50	-47	-0.0953	-7470A	PPB	06 Dec 2023 12:55:27	U	SMPL
CRA - 1	CRA	1285	0.2043	-7470A	PPB	06 Dec 2023 13:00:19	U	SMPL
HighStd - 1	HighStd	42306	9.4293	-7470A	PPB	06 Dec 2023 13:02:34	U	SMPL
ChkStd - 1	ChkStd	31241	6.9409	-7470A	PPB	06 Dec 2023 13:04:49	U	SMPL
PB157587BS - 1	PB157587BS	18211	4.0107	-7470A	PPB	06 Dec 2023 13:09:26	U	SMPL
O5588-03 - 1	TUDOR	-223	-0.1349	-7470A	PPB	06 Dec 2023 13:11:43	U	SMPL
O5602-02 - 1	SPRING-VALLEY-TP-1	0	-0.0847	-7470A	PPB	06 Dec 2023 13:14:01	U	SMPL
O5602-04 - 1	SPRING-VALLEY-TP-2	20	-0.0802	-7470A	PPB	06 Dec 2023 13:16:18	U	SMPL
O5602-06 - 1	SPRING-VALLEY-TP-3	106	-0.0609	-7470A	PPB	06 Dec 2023 13:18:35	U	SMPL
O5602-08 - 1	ROUTE-4-TP-4	75	-0.0679	-7470A	PPB	06 Dec 2023 13:20:50	U	SMPL
CCV51 - 1	CCV51	22817	5.0465	-7470A	PPB	06 Dec 2023 13:23:06	U	SMPL
CCB51 - 1	CCB51	108	-0.0604	-7470A	PPB	06 Dec 2023 13:25:21	U	SMPL
O5602-10 - 1	ROUTE-4-TP-5	0	-0.0847	-7470A	PPB	06 Dec 2023 13:27:39	U	SMPL
O5602-12 - 1	ROUTE-4-TP-6	41	-0.0755	-7470A	PPB	06 Dec 2023 13:29:54	U	SMPL
O5602-12DUP - 1	ROUTE-4-TP-6DUP	53	-0.0728	-7470A	PPB	06 Dec 2023 13:32:10	U	SMPL
O5602-12MS - 1	ROUTE-4-TP-6MS	18880	4.1611	-7470A	PPB	06 Dec 2023 13:34:26	U	SMPL
O5602-12MSD - 1	ROUTE-4-TP-6MSD	16782	3.6893	-7470A	PPB	06 Dec 2023 13:36:42	U	SMPL
PB157587BL - 1	PBW	40	-0.0757	-7470A	PPB	06 Dec 2023 13:39:01	U	SMPL
PB157520TB - 1	PB157520TB	-28	-0.0910	-7470A	PPB	06 Dec 2023 13:41:22	U	SMPL
O5602-12LX5 - 1		21	-0.0800	-7470A	PPB	06 Dec 2023 13:43:39	U	SMPL
O5602-12A - 1		18996	4.1872	-7470A	PPB	06 Dec 2023 13:45:57	U	SMPL
CCV52 - 1	CCV52	22529	4.9817	-7470A	PPB	06 Dec 2023 13:48:12	U	SMPL
CCB52 - 1	CCB52	255	-0.0274	-7470A	PPB	06 Dec 2023 13:53:01	U	SMPL