



Analytical Summary Report

Analysis Method: 7196A

ANALYST: Rubina

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: Iwona

Run Number: LB128579

pH Meter ID: WC pH Meter-1

Reagent/Standard	Lot/Log #
hexavalent chromium color reagent	WP105503
5N sulfuric acid	WP105081
HNO3 Hex-Chrome, 5M	WP105080
Hexchrome Cleaning Solution	WP103556

Intercept: 0.0022

Slope: 0.7659

Regression: 0.999993

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml)	Final Vol (ml)	pH HNO3	pH H2SO4	Absorb.at 540nm		Absorbance Difference	Result (mg/L)	%D	Anal Date	Anal Time
								Backgrnd	Color					
1	CAL1	0	1	100	100	7.28	1.72	0.000	0.000	0.000	-0.00		12/06/2023	12:00
2	CAL2	0.01	1	100	100	7.36	1.84	0.000	0.010	0.010	0.010	0	12/06/2023	12:01
3	CAL3	0.025	1	100	100	7.38	1.87	0.000	0.022	0.022	0.025	0	12/06/2023	12:02
4	CAL4	0.05	1	100	100	7.35	1.85	0.000	0.041	0.041	0.050	0	12/06/2023	12:03
5	CAL5	0.1	1	100	100	7.38	1.88	0.000	0.080	0.080	0.101	1	12/06/2023	12:04
6	CAL6	0.5	1	100	100	7.40	1.91	0.000	0.385	0.385	0.499	-0.2	12/06/2023	12:05
7	CAL7	1	1	100	100	7.36	1.84	0.000	0.768	0.768	0.999	-0.1	12/06/2023	12:06



Analytical Summary Report

Analysis Method: 7196A

ANALYST: Rubina

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: Iwona

Run Number: LB128579

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb.at540nm		Absorbance Difference	Intermediate Result (mg/L)	Anal Date	Anal Time
								Backgrnd	Color				
1	ICV	0.5	1	100	100	7.41	1.88	0.000	0.386	0.386	0.501	12/06/2023	12:07
2	ICB		1	100	100	7.25	1.76	0.000	0.001	0.001	-0.002	12/06/2023	12:08
3	CCV1	0.5	1	100	100	7.42	1.95	0.000	0.383	0.383	0.497	12/06/2023	12:09
4	CCB1		1	100	100	7.24	1.77	0.000	0.001	0.001	-0.002	12/06/2023	12:10
5	RL Check	0.01	1	100	100	7.42	1.92	0.000	0.010	0.010	0.010	12/06/2023	12:11
6	PB157567BL		1	2.50	100	7.28	1.74	0.000	0.001	0.001	-0.002	12/06/2023	12:12
7	PB157567BS	20	1	2.50	100	7.42	1.91	0.000	0.385	0.385	0.500	12/06/2023	12:13
8	O5602-01		1	2.56	100	7.52	2.08	0.038	0.038	0.000	-0.003	12/06/2023	12:14
9	O5602-03		1	2.54	100	7.62	2.16	0.030	0.032	0.002	0.000	12/06/2023	12:15
10	O5602-05		1	2.53	100	7.56	2.06	0.004	0.004	0.000	-0.003	12/06/2023	12:16
11	O5602-07		1	2.53	100	7.58	2.10	0.010	0.010	0.000	-0.003	12/06/2023	12:17
12	O5602-09		1	2.55	100	7.52	2.14	0.032	0.032	0.000	-0.003	12/06/2023	12:18
13	O5602-11		1	2.52	100	7.52	2.08	0.032	0.032	0.000	-0.003	12/06/2023	12:19
14	O5629-01		1	2.54	100	7.54	2.06	0.004	0.006	0.002	0.000	12/06/2023	12:20
15	O5629-01DU		1	2.54	100	7.52	2.08	0.005	0.006	0.001	-0.002	12/06/2023	12:21
16	CCV2	0.5	1	100	100	7.42	1.93	0.000	0.383	0.383	0.497	12/06/2023	12:22
17	CCB2		1	100	100	7.26	1.74	0.000	0.001	0.001	-0.002	12/06/2023	12:23
18	O5629-01MS	40	2	2.55	100	7.56	2.06	0.003	0.359	0.356	0.462	12/06/2023	12:24
19	O5629-01MS	1284	40	2.55	100	7.52	2.08	0.000	0.612	0.612	0.796	12/06/2023	12:25
20	O5629-01MS	40	2	2.53	100	7.54	2.08	0.003	0.383	0.380	0.493	12/06/2023	12:26
21	O5629-04		1	2.55	100	7.58	2.10	0.012	0.014	0.002	0.000	12/06/2023	12:27
22	O5630-01		1	2.55	100	7.64	2.08	0.012	0.014	0.002	0.000	12/06/2023	12:28
23	O5630-03		1	2.52	100	7.62	2.18	0.010	0.012	0.002	0.000	12/06/2023	12:29
24	O5636-01		1	2.56	100	7.66	2.10	0.006	0.008	0.002	0.000	12/06/2023	12:30
25	CCV3	0.5	1	100	100	7.44	1.94	0.000	0.385	0.385	0.500	12/06/2023	12:31
26	CCB3		1	100	100	7.25	1.78	0.000	0.001	0.001	-0.002	12/06/2023	12:32