



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB128580

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	12/06/2023	12:30
2	CAL2	1	Water	NA	NA	20.4	7.00	12/06/2023	12:31
3	CAL3	1	Water	NA	NA	20.5	10.02	12/06/2023	12:35
4	ICV	1	Water	NA	NA	20.4	6.99	12/06/2023	12:37
5	CCV1	1	Water	NA	NA	20.5	2.01	12/06/2023	12:40
6	O5640-01	1	Water	NA	NA	21.9	5.43	12/06/2023	12:50
7	O5640-01DUP	1	Water	NA	NA	21.8	5.44	12/06/2023	12:51
8	O5641-01	1	Water	NA	NA	21.6	5.54	12/06/2023	13:00
9	O5641-02	1	Water	NA	NA	22.4	5.65	12/06/2023	13:10
10	CCV2	1	Water	NA	NA	20.5	12.02	12/06/2023	13:15

UB128580

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O5640 WorkList ID : 176094 Department : Wet-Chemistry Date : 12-06-2023 12:07:11

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5640-01	A GAMMA-JARINU	Water	pH	Cool 4 deg C	STER18	L11	11/28/2023	9040C
O5641-01	A JS9600	Water	pH	Cool 4 deg C	STER18	L11	11/28/2023	9040C
O5641-02	A JS7500	Water	pH	Cool 4 deg C	STER18	L11	11/28/2023	9040C

Date/Time 12106123 12:20
Raw Sample Received by: JB (QC)
Raw Sample Relinquished by: AS

Date/Time 12106123 14:00
Raw Sample Received by: AS
Raw Sample Relinquished by: JB (QC)