



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB128585

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/06/2023	14:10
2	CAL2	1	Water	NA	NA	20.3	7.00	12/06/2023	14:11
3	CAL3	1	Water	NA	NA	20.4	10.02	12/06/2023	14:12
4	ICV	1	Water	NA	NA	20.5	7.01	12/06/2023	14:15
5	CCV1	1	Water	NA	NA	20.5	2.01	12/06/2023	14:20
6	O5650-01	1	Water	NA	NA	20.4	5.69	12/06/2023	14:25
7	O5650-01DUP	1	Water	NA	NA	20.5	5.70	12/06/2023	14:26
8	O5650-02	1	Water	NA	NA	20.7	5.93	12/06/2023	14:35
9	O5650-03	1	Water	NA	NA	20.2	5.43	12/06/2023	14:40
10	O5650-04	1	Water	NA	NA	20.4	8.47	12/06/2023	14:50
11	CCV2	1	Water	NA	NA	20.3	12.02	12/06/2023	14:52

VB 128585

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 05650

WorkList ID : 176098

Department : Wet-Chemistry

Date : 12-06-2023 13:43:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5650-01	C 50194	Water	pH	Cool 4 deg C	PSEG03	L31	12/06/2023	9040C
O5650-02	C 50195	Water	pH	Cool 4 deg C	PSEG03	L31	12/06/2023	9040C
O5650-03	C 50196	Water	pH	Cool 4 deg C	PSEG03	L31	12/06/2023	9040C
O5650-04	D 50197	Water	pH	Cool 4 deg C	PSEG03	L31	12/06/2023	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: