



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB128591

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	12/06/2023	16:45
2	CAL2	1	Water	NA	NA	20.5	7.00	12/06/2023	16:46
3	CAL3	1	Water	NA	NA	20.5	10.03	12/06/2023	16:47
4	ICV	1	Water	NA	NA	20.5	7.01	12/06/2023	16:50
5	CCV1	1	Water	NA	NA	20.5	2.01	12/06/2023	17:00
6	O5628-02	1	Solid	20.02	20	20.4	7.64	12/06/2023	17:02
7	O5628-04	1	Solid	20.03	20	20.4	8.70	12/06/2023	17:05
8	O5628-06	1	Solid	20.04	20	20.3	9.42	12/06/2023	17:10
9	O5628-08	1	Solid	20.03	20	20.2	8.44	12/06/2023	17:15
10	O5639-02	1	Solid	20.02	20	23.6	9.01	12/06/2023	17:20
11	O5657-02	1	Solid	20.02	20	22.3	7.19	12/06/2023	17:25
12	O5671-04	1	Solid	20.03	20	20.3	8.34	12/06/2023	17:30
13	O5671-08	1	Solid	20.05	20	20.3	8.45	12/06/2023	17:35
14	O5671-08DUP	1	Solid	20.04	20	20.4	8.46	12/06/2023	17:36
15	CCV2	1	Water	NA	NA	20.3	12.02	12/06/2023	17:39

WORKLIST(Hardcopy Internal Chain)

128591

WorkList Name : corosovity O5639 WorkList ID : 176097 Department : Wet-Chemistry Date : 12-06-2023 12:29:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5628-02	SB-01-COMP	Solid	Corrosivity	Cool 4 deg C	LIRO01	L11	12/04/2023	9045D
O5628-04	SB-02-COMP	Solid	Corrosivity	Cool 4 deg C	LIRO01	L11	12/04/2023	9045D
O5628-06	SB-04-COMP	Solid	Corrosivity	Cool 4 deg C	LIRO01	L11	12/04/2023	9045D
O5628-08	SB-03-COMP	Solid	Corrosivity	Cool 4 deg C	LIRO01	L11	12/04/2023	9045D
O5639-02	E8965C-CC-001	Solid	Corrosivity	Cool 4 deg C	ENTA05	L41	12/05/2023	9045D
O5657-02	205-199	Solid	Corrosivity	Cool 4 deg C	PSEG03	L21	12/06/2023	9045D
O5671-04	TP-34	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	12/06/2023	9045D
O5671-08	TP-35	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	12/06/2023	9045D

Date/Time 12/06/23 16:40
 Raw Sample Received by: J01 WDC
 Raw Sample Relinquished by: CDS

Date/Time 12/06/23 18:30
 Raw Sample Received by: CDS
 Raw Sample Relinquished by: J01 WDC