

LB128597


7471B

INSTRUMENT MDI CV1

Linear

 μ Abs.:

65361



A= 0.0000e+000

B= 1.5508e-004

C= -3.1567e-002

Rho= 0.9996721

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Std
0.0	0.000	0.029	0.029	389	0.000	389					0.00
0.2	0.200	0.234	0.034	1713	0.0 %	1713					17
2.5	2.500	2.433	-0.067	15893	0.0 %	15893					-3
5.0	5.000	5.074	0.074	32922	0.0 %	32922					1
7.5	7.500	7.325	-0.175	47436	0.0 %	47436					-2
10.0	10.000	10.105	0.105	65361	0.0 %	65361					1

LB128597 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 07 Dec 2023 09:28:59

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SC	389	-	0.0000	7471B	PPB	07 Dec 2023 09:44:46	S	Std
0.2 - 1	SC.2	1713	-	0.2000	7471B	PPB	07 Dec 2023 09:47:03	S	Std
2.5 - 1	12.5	15893	-	2.5000	7471B	PPB	07 Dec 2023 09:49:20	S	Std
5.0 - 1	5.5	32922	-	5.0000	7471B	PPB	07 Dec 2023 09:51:37	S	Std
7.5 - 1	57.5	47436	-	7.5000	7471B	PPB	07 Dec 2023 09:53:57	S	Std
10.0 - 1	516	65361	-	10.0000	7471B	PPB	07 Dec 2023 09:56:23	S	Std
ICV06 - 1	ICV06	25257	3.8854	-	7471B	PPB	07 Dec 2023 10:02:15	U	SMPL
ICB06 - 1	ICB06	-268	-0.0731	-	7471B	PPB	07 Dec 2023 10:04:30	U	SMPL
CCV53 - 1	CCV53	32829	5.0597	-	7471B	PPB	07 Dec 2023 10:06:51	U	SMPL
CCB53 - 1	CCB53	-181	-0.0596	-	7471B	PPB	07 Dec 2023 10:09:05	U	SMPL
CRA - 1	CRA	1584	0.2141	-	7471B	PPB	07 Dec 2023 10:11:31	U	SMPL
HighStd - 1	HighStd	58925	9.1068	-	7471B	PPB	07 Dec 2023 10:13:46	U	SMPL
ChkStd - 1	ChkStd	43270	6.6789	-	7471B	PPB	07 Dec 2023 10:18:51	U	SMPL
PB157591BL - 1	PBS	-368	-0.0886	-	7471B	PPB	07 Dec 2023 10:21:10	U	SMPL
PB157591BS - 1	LCSS	22661	3.4828	-	7471B	PPB	07 Dec 2023 10:26:15	U	SMPL
O5047-10 - 1	PT-MET-SOIL	277223	42.9612	-	7471B	PPB	07 Dec 2023 10:28:33	U	SMPL
O5616-04 - 1	1150	1721	0.2353	-	7471B	PPB	07 Dec 2023 10:30:58	U	SMPL
O5617-01 - 1	HD-02-120423	2422	0.3440	-	7471B	PPB	07 Dec 2023 10:33:55	U	SMPL
O5624-01 - 1	14TH-SOIL-1	1897	0.2626	-	7471B	PPB	07 Dec 2023 10:36:14	U	SMPL
O5624-03 - 1	14TH-SOIL-2	1302	0.1704	-	7471B	PPB	07 Dec 2023 10:38:31	U	SMPL
CCV54 - 1	CCV54	30137	4.6422	-	7471B	PPB	07 Dec 2023 10:40:47	U	SMPL
CCB54 - 1	CCB54	-105	-0.0479	-	7471B	PPB	07 Dec 2023 10:43:03	U	SMPL
O5629-01 - 1	VNJ-216	906	0.1089	-	7471B	PPB	07 Dec 2023 10:45:30	U	SMPL
O5629-04 - 1	72-11985	565	0.0561	-	7471B	PPB	07 Dec 2023 10:47:46	U	SMPL
O5630-01 - 1	EO-01-120523	3825	0.5616	-	7471B	PPB	07 Dec 2023 10:50:02	U	SMPL
O5630-03 - 1	EO-02-120523	3773	0.5536	-	7471B	PPB	07 Dec 2023 10:52:17	U	SMPL
O5636-01 - 1	TR-05-120523	7597	1.1466	-	7471B	PPB	07 Dec 2023 10:54:35	U	SMPL
O5636-01DUP - 1	TR-05-120523DUP	6877	1.0349	-	7471B	PPB	07 Dec 2023 10:56:53	U	SMPL
O5636-01MS - 1	TR-05-120523MS	31871	4.9111	-	7471B	PPB	07 Dec 2023 10:59:12	U	SMPL
O5636-01MSD - 1	TR-05-120523MSD	32967	5.0811	-	7471B	PPB	07 Dec 2023 11:01:31	U	SMPL
O5636-01LX5 - 1		1592	0.2153	-	7471B	PPB	07 Dec 2023 11:25:50	U	SMPL
O5636-01A - 1		29820	4.5930	-	7471B	PPB	07 Dec 2023 11:28:05	U	SMPL
CCV55 - 1	CCV55	29958	4.6144	-	7471B	PPB	07 Dec 2023 11:30:22	U	SMPL
CCB55 - 1	CCB55	-64	-0.0415	-	7471B	PPB	07 Dec 2023 11:32:47	U	SMPL
O5047-10DLX10 - 1	PT-MET-SOIL	31891	4.9142	-	7471B	PPB	07 Dec 2023 11:37:28	U	SMPL
CCV56 - 1	CCV56	31790	4.8985	-	7471B	PPB	07 Dec 2023 11:45:03	U	SMPL
CCB56 - 1	CCB56	-91	-0.0457	-	7471B	PPB	07 Dec 2023 11:47:21	U	SMPL