



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128600

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	12/07/2023	08:15
2	CAL2	1	Water	NA	NA	20.6	7.00	12/07/2023	08:16
3	CAL3	1	Water	NA	NA	20.5	10.03	12/07/2023	08:20
4	ICV	1	Water	NA	NA	20.5	6.98	12/07/2023	08:25
5	CCV1	1	Water	NA	NA	20.5	2.01	12/07/2023	08:30
6	O5658-01	1	Solid	20.02	20	23.2	6.43	12/07/2023	08:35
7	O5658-01DUP	1	Solid	20.03	20	23.3	6.44	12/07/2023	08:36
8	CCV2	1	Water	NA	NA	20.6	12.02	12/07/2023	08:40

W3 128600

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 05671		WorkList ID : 176105		Department : Wet-Chemistry		Date : 12-07-2023 07:52:45	
Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date Method
O5658-01	60119	Solid	pH	Cool 4 deg C	PSEG03	L21	12/06/2023 9045D

Date/Time 12/07/23 08:10
Raw Sample Received by: JHCOU
Raw Sample Relinquished by: JHCOU

Date/Time 12/07/23 11:00
Raw Sample Received by: JHCOU
Raw Sample Relinquished by: JHCOU