



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB128606

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/07/2023	08:25
2	CAL2	1	Water	NA	NA	20.3	7.00	12/07/2023	08:26
3	CAL3	1	Water	NA	NA	20.3	10.02	12/07/2023	08:29
4	ICV	1	Water	NA	NA	20.6	7.01	12/07/2023	08:33
5	CCV1	1	Water	NA	NA	20.5	2.01	12/07/2023	08:35
6	O5658-02	1	Water	NA	NA	22.3	7.50	12/07/2023	08:40
7	O5658-02DUP	1	Water	NA	NA	22.4	7.51	12/07/2023	08:41
8	O5658-03	1	Water	NA	NA	20.3	6.06	12/07/2023	08:45
9	O5672-01	1	Water	NA	NA	20.1	5.82	12/07/2023	08:50
10	O5672-03	1	Water	NA	NA	20.2	6.99	12/07/2023	08:55
11	O5672-05	1	Water	NA	NA	20.6	6.85	12/07/2023	09:00
12	O5672-07	1	Water	NA	NA	20.2	6.77	12/07/2023	09:05
13	CCV2	1	Water	NA	NA	20.5	12.02	12/07/2023	09:10

VB 128606

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 05656 water		WorkList ID : 176106		Department : Wet-Chemistry		Date : 12-07-2023 07:55:03		
Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5658-02	A 34147	Water	pH	Cool 4 deg C	PSEG03	L21	12/06/2023	9040C
O5658-03	C 60118	Water	pH	Cool 4 deg C	PSEG03	L21	12/06/2023	9040C
O5672-01	CA MW-1	Water	pH	Cool 4 deg C	LOCK01	L31	12/05/2023	9040C
O5672-03	CG MW-3	Water	pH	Cool 4 deg C	LOCK01	L31	12/05/2023	9040C
O5672-05	CG MW-4	Water	pH	Cool 4 deg C	LOCK01	L31	12/05/2023	9040C
O5672-07	CG MW-2	Water	pH	Cool 4 deg C	LOCK01	L31	12/05/2023	9040C

Date/Time 12/07/23 08:20
 Raw Sample Received by: JG CWC
 Raw Sample Relinquished by: RM SM

Date/Time 12/07/23 11:00
 Raw Sample Received by: RM SM
 Raw Sample Relinquished by: JG CWC