

7470A

Linear

 $\mu \text{ Abs.} :$

45094

A= 0.0000e+000

B= 2.2473e-004

C= -2.4980e-002

Rho= 0.9998000

Accept=Accepted

Max. Conc:— 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	avg
0.0	0.000	0.015	0.015	179	0.000	179	0				179
0.2	0.200	0.233	0.033	1147	0.0 %	1147					1147
2.5	2.500	2.499	-0.001	11230	0.0 %	11230					11230
5.0	5.000	4.983	-0.017	22285	0.0 %	22285					22285
7.5	7.500	7.361	-0.139	32868	0.0 %	32868					32868
10.0	10.000	10.109	0.109	45094	0.0 %	45094					45094

LB128610 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Dec 2023 13:06:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	5	179	-	0.0000	7470A	PPB	07 Dec 2023 13:13:26	S	Std
0.2 - 1	50.2	1147	-	0.2000	7470A	PPB	07 Dec 2023 13:15:43	S	Std
2.5 - 1	52.5	11230	-	2.5000	7470A	PPB	07 Dec 2023 13:18:00	S	Std
5.0 - 1	55	22285	-	5.0000	7470A	PPB	07 Dec 2023 13:20:17	S	Std
7.5 - 1	53	32868	-	7.5000	7470A	PPB	07 Dec 2023 13:22:35	S	Std
10.0 - 1	52	45094	-	10.0000	7470A	PPB	07 Dec 2023 13:27:33	S	Std
ICV07 - 1	ICV07	16956	3.7855	-	7470A	PPB	07 Dec 2023 13:31:07	U	SMPL
ICB07 - 1	ICB07	-129	-0.0540	-	7470A	PPB	07 Dec 2023 13:33:22	U	SMPL
CCV57 - 1	CCV57	22846	5.1091	-	7470A	PPB	07 Dec 2023 13:35:39	U	SMPL
CCB57 - 1	CCB57	-159	-0.0607	-	7470A	PPB	07 Dec 2023 13:37:54	U	SMPL
CRA - 1	CRA	1088	0.2195	-	7470A	PPB	07 Dec 2023 13:40:12	U	SMPL
HighStd - 1	HighStd	45661	10.2363	-	7470A	PPB	07 Dec 2023 13:42:27	U	SMPL
ChkStd - 1	ChkStd	31004	6.9425	-	7470A	PPB	07 Dec 2023 13:44:43	U	SMPL
PB157601BL - 1	PBW	-249	-0.0809	-	7470A	PPB	07 Dec 2023 13:47:01	U	SMPL
PB157601BS - 1	LCSW	16403	3.6612	-	7470A	PPB	07 Dec 2023 13:49:20	U	SMPL
O5612-01 - 1	MW-06I	-92	-0.0457	-	7470A	PPB	07 Dec 2023 13:51:36	U	SMPL
O5612-01DUP - 1	MW-06IDUP	144	0.0074	-	7470A	PPB	07 Dec 2023 13:53:55	U	SMPL
O5612-02 - 1	O5612-01MS	14026	3.1270	-	7470A	PPB	07 Dec 2023 13:56:12	U	SMPL
O5612-03 - 1	O5612-01MSD	14224	3.1715	-	7470A	PPB	07 Dec 2023 13:58:29	U	SMPL
O5612-04 - 1	MW-04I	1859	0.3928	-	7470A	PPB	07 Dec 2023 14:00:46	U	SMPL
CCV58 - 1	CCV58	22151	4.9529	-	7470A	PPB	07 Dec 2023 14:03:04	U	SMPL
CCB58 - 1	CCB58	-155	-0.0598	-	7470A	PPB	07 Dec 2023 14:05:22	U	SMPL
O5612-05 - 1	MW-06S	133	0.0049	-	7470A	PPB	07 Dec 2023 14:07:40	U	SMPL
O5612-07 - 1	MW-06I-DUP	258	0.0330	-	7470A	PPB	07 Dec 2023 14:09:55	U	SMPL
O5616-02 - 1	1149	886	0.1741	-	7470A	PPB	07 Dec 2023 14:12:10	U	SMPL
O5620-01 - 1	OILY-WATER-TOTE	2368	0.5072	-	7470A	PPB	07 Dec 2023 14:14:26	U	SMPL
O5650-04 - 1	50197	156	0.0101	-	7470A	PPB	07 Dec 2023 14:16:44	U	SMPL
O5658-03 - 1	60118	2318	0.4959	-	7470A	PPB	07 Dec 2023 14:19:02	U	SMPL
O5660-02 - 1	1100	2284	0.4883	-	7470A	PPB	07 Dec 2023 14:21:18	U	SMPL
O5612-01LX5 - 1		-122	-0.0524	-	7470A	PPB	07 Dec 2023 14:23:37	U	SMPL
O5612-01A - 1		16244	3.6255	-	7470A	PPB	07 Dec 2023 14:25:56	U	SMPL
CCV59 - 1	CCV59	22529	5.0379	-	7470A	PPB	07 Dec 2023 14:28:12	U	SMPL
CCB59 - 1	CCB59	-81	-0.0432	-	7470A	PPB	07 Dec 2023 14:30:30	U	SMPL