



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB128612

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP105606
Turbidity Calibration std, 20NTU	WP105601
Turbidity Calibration std, 40NTU	WP105600
Turbidity Calibration std, 1NTU	WP105604
Turbidity Calibration std, 5NTU	WP105603
Turbidity Calibration - CCV std, 10 NTU	WP105602
10 NTU Standard 500 ml	W2971

Intercept: 0.1323

Slope: 0.9824

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.146	0.01		12/07/2023	09:40
2	CAL2	1	1	1.230	1.12	11.7	12/07/2023	09:43
3	CAL3	5	1	4.680	4.63	-7.4	12/07/2023	09:46
4	CAL4	10	1	10.200	10.25	2.5	12/07/2023	09:50
5	CAL5	20	1	19.800	20.02	0.1	12/07/2023	09:53
6	CAL6	40	1	39.400	39.97	-0.1	12/07/2023	09:57



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB128612

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.600	10.655	12/07/2023	10:00
2	ICB		1	0.155	0.023	12/07/2023	10:03
3	CCV1	10	1	10.500	10.553	12/07/2023	10:06
4	CCB1		1	0.157	0.025	12/07/2023	10:09
5	LB128612BL		1	0.154	0.022	12/07/2023	10:12
6	O5672-01		1	1.450	1.341	12/07/2023	10:15
7	O5672-01DUP		1	1.420	1.311	12/07/2023	10:18
8	O5672-03		1	1.290	1.178	12/07/2023	10:21
9	O5672-05		1	1.320	1.209	12/07/2023	10:24
10	O5672-07		1	11.900	11.979	12/07/2023	10:27
11	CCV2	10	1	10.100	10.146	12/07/2023	10:30
12	CCB2		1	0.156	0.024	12/07/2023	10:33

LB128612

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TURB-120723 WorkList ID : 176119 Department : Wet-Chemistry Date : 12-07-2023 09:11:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5672-01	MW-1	Water	Turbidity	Cool 4 deg C	LOCK01	L31	12/05/2023	SM2130 B
O5672-03	MW-3	Water	Turbidity	Cool 4 deg C	LOCK01	L31	12/05/2023	SM2130 B
O5672-05	MW-4	Water	Turbidity	Cool 4 deg C	LOCK01	L31	12/05/2023	SM2130 B
O5672-07	MW-2	Water	Turbidity	Cool 4 deg C	LOCK01	L31	12/05/2023	SM2130 B

Date/Time 12/07/23 09:15
Raw Sample Received by: 12/07/23
Raw Sample Relinquished by: 12/07/23

Date/Time 12/07/23 11:20
Raw Sample Received by: 12/07/23
Raw Sample Relinquished by: 12/07/23