



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB128622

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	12/07/2023	16:15
2	CAL2	1	Water	NA	NA	20.6	7.00	12/07/2023	16:16
3	CAL3	1	Water	NA	NA	20.6	10.02	12/07/2023	16:19
4	ICV	1	Water	NA	NA	20.5	7.01	12/07/2023	16:20
5	CCV1	1	Water	NA	NA	20.5	2.01	12/07/2023	16:25
6	O5691-03	1	Water	NA	NA	20.3	7.06	12/07/2023	16:30
7	O5700-01	1	Water	NA	NA	20.1	6.87	12/07/2023	16:35
8	O5702-01	1	Water	NA	NA	20.3	7.15	12/07/2023	16:40
9	O5702-01DUP	1	Water	NA	NA	20.4	7.16	12/07/2023	16:41
10	CCV2	1	Water	NA	NA	20.5	12.02	12/07/2023	16:45

WORKLIST(Hardcopy Internal Chain)

VB 128622

WorkList Name : PH O5700 WorkList ID : 176131 Department : Wet-Chemistry Date : 12-07-2023 14:41:17

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5691-03	D A3909	Water	pH	Cool 4 deg C	PSEG03	L21	12/07/2023	9040C
O5700-01	A 257419	Water	pH	Cool 4 deg C	PSEG03	L31	12/07/2023	9040C
O5702-01	A POOL-WATER	Water	pH	Cool 4 deg C	STER01	L11	12/05/2023	9040C

Date/Time 12/07/23 16:10
 Raw Sample Received by: J060C1
 Raw Sample Relinquished by: JTC5M

Date/Time 12/07/23 18:00
 Raw Sample Received by: JTC5M
 Raw Sample Relinquished by: J060C1