

LB128629


SFAM01.1

INSTRUMENT ID : CV1

Linear

μ Abs.:

45977



A= 0.0000e+000

B= 2.2049e-004 slope

C= -1.6013e-001 y-intercept

Rho= 0.9999170

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% of 1
0.00	0.000	-0.061	-0.061	448	0.000	448					1
0.05	0.050					442	0				3
0.20	0.200	0.195	-0.005	1612	0.0 %	1612					4
2.50	2.500	2.596	0.096	12498	0.0 %	12498					0
5.00	5.000	4.989	-0.011	23355	0.0 %	23355					0
7.50	7.500	7.504	0.004	34757	0.0 %	34757					0
10.0	10.000	9.977	-0.023	45977	0.0 %	45977					0

LB128623

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	448	-		0 SFAM01.1	PPB	12/7/2023 16:35	Std
	0.2 S01	1612	-		0.2 SFAM01.1	PPB	12/7/2023 16:37	Std
	2.5 S02	12498	-		2.5 SFAM01.1	PPB	12/7/2023 16:39	Std
	5 S03	23355	-		5 SFAM01.1	PPB	12/7/2023 16:42	Std
	7.5 S04	34757	-		7.5 SFAM01.1	PPB	12/7/2023 16:44	Std
	10 S05	45977	-		10 SFAM01.1	PPB	12/7/2023 16:46	Std
ICV008	ICV008	19649	4.1723 -		SFAM01.1	PPB	12/7/2023 16:49	SMPL
ICB008	ICB008	693	-0.0073 -		SFAM01.1	PPB	12/7/2023 16:54	SMPL
CCV060	CCV060	23396	4.9985 -		SFAM01.1	PPB	12/7/2023 16:56	SMPL
CCB060	CCB060	483	-0.0536 -		SFAM01.1	PPB	12/7/2023 16:59	SMPL
PB157603BL	pbs603	394	-0.0733 -		SFAM01.1	PPB	12/7/2023 17:01	SMPL
O5493-01	MEZZC1	65534	14.2896 -		SFAM01.1	PPB	12/7/2023 17:03	SMPL
O5493-02	MEZZC2	74903	16.3554 -		SFAM01.1	PPB	12/7/2023 17:05	SMPL
O5493-03	MEZZC3	62755	13.6769 -		SFAM01.1	PPB	12/7/2023 17:08	SMPL
O5493-04	MEZZC3D	58039	12.637 -		SFAM01.1	PPB	12/7/2023 17:10	SMPL
O5493-05	MEZZC3S	71873	15.6873 -		SFAM01.1	PPB	12/7/2023 17:13	SMPL
O5493-01DLX5	MEZZC1	12618	2.622 -		SFAM01.1	PPB	12/7/2023 17:15	SMPL
O5493-02DLX5	MEZZC2	13805	2.8838 -		SFAM01.1	PPB	12/7/2023 17:17	SMPL
O5493-03DLX5	MEZZC3	12672	2.634 -		SFAM01.1	PPB	12/7/2023 17:20	SMPL
O5493-04DLX5	MEZZC3D	12671	2.6337 -		SFAM01.1	PPB	12/7/2023 17:22	SMPL
O5493-05DLX5	MEZZC3S	14619	3.0633 -		SFAM01.1	PPB	12/7/2023 17:24	SMPL
CCV061	CCV061	23450	5.0104 -		SFAM01.1	PPB	12/7/2023 17:27	SMPL
CCB061	CCB061	475	-0.0554 -		SFAM01.1	PPB	12/7/2023 17:29	SMPL