

LB128625


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44482



A= 0.0000e+000

B= 2.2455e-004 slope

C= 2.8008e-002 y -intercept

Rho= 0.9984855

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.050	0.050	97	0.000	97					
0.05	0.050					442	0				
0.20	0.200	0.235	0.035	920	0.0 %	920					
2.50	2.500	2.649	0.149	11672	0.0 %	11672					
5.00	5.000	4.567	-0.433	20214	0.0 %	20214					
7.50	7.500	7.683	0.183	34093	0.0 %	34093					
10.0	10.000	10.016	0.016	44482	0.0 %	44482					

0.01
= 18
6
9
2
0

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	97	-		0 SFAM01.1	PPB	12/7/2023 18:28	Std
	0.2 S01	920	-		0.2 SFAM01.1	PPB	12/7/2023 18:30	Std
	2.5 S02	11672	-		2.5 SFAM01.1	PPB	12/7/2023 18:33	Std
	5 S03	20214	-		5 SFAM01.1	PPB	12/7/2023 18:35	Std
	7.5 S04	34093	-		7.5 SFAM01.1	PPB	12/7/2023 18:37	Std
	10 S05	44482	-		10 SFAM01.1	PPB	12/7/2023 18:39	Std
ICV011	ICV011	16986	3.8421 -		SFAM01.1	PPB	12/7/2023 18:43	SMPL
ICB011	ICB011	-229	-0.0234 -		SFAM01.1	PPB	12/7/2023 18:45	SMPL
CCV066	CCV066	20576	4.6483 -		SFAM01.1	PPB	12/7/2023 18:50	SMPL
CCB066	CCB066	-257	-0.0297 -		SFAM01.1	PPB	12/7/2023 18:52	SMPL
PB157631BL	PBW631	5	0.0291 -		SFAM01.1	PPB	12/7/2023 18:54	SMPL
O5494-01	MC0AA1	4	0.0289 -		SFAM01.1	PPB	12/7/2023 18:57	SMPL
O5494-02	MC0AC8	35	0.0359 -		SFAM01.1	PPB	12/7/2023 18:59	SMPL
O5494-03	MC0AC9	46	0.0383 -		SFAM01.1	PPB	12/7/2023 19:01	SMPL
O5494-04	MC0AC9D	65	0.0426 -		SFAM01.1	PPB	12/7/2023 19:03	SMPL
O5494-05	MC0AC9S	4645	1.071 -		SFAM01.1	PPB	12/7/2023 19:06	SMPL
O5494-06	MC0AD0	246	0.0832 -		SFAM01.1	PPB	12/7/2023 19:08	SMPL
O5494-07	MC0AD1	156	0.063 -		SFAM01.1	PPB	12/7/2023 19:10	SMPL
O5494-08	MC0AD2	94	0.0491 -		SFAM01.1	PPB	12/7/2023 19:12	SMPL
O5494-09	MC0AD3	101	0.0507 -		SFAM01.1	PPB	12/7/2023 19:15	SMPL
O5494-10	MC0AD4	106	0.0518 -		SFAM01.1	PPB	12/7/2023 19:17	SMPL
O5494-11	MC0AD5	110	0.0527 -		SFAM01.1	PPB	12/7/2023 19:19	SMPL
O5494-12	MC0AD6	133	0.0579 -		SFAM01.1	PPB	12/7/2023 19:21	SMPL
O5494-13	MC0AD7	174	0.0671 -		SFAM01.1	PPB	12/7/2023 19:24	SMPL
O5494-14	MC0AD8	88	0.0478 -		SFAM01.1	PPB	12/7/2023 19:26	SMPL
O5494-15	MC0AD9	127	0.0565 -		SFAM01.1	PPB	12/7/2023 19:28	SMPL
O5494-16	MC0AE0	169	0.066 -		SFAM01.1	PPB	12/7/2023 19:30	SMPL
O5494-17	MC0AE1	162	0.0644 -		SFAM01.1	PPB	12/7/2023 19:33	SMPL
O5494-18	MC0AE4	122	0.0554 -		SFAM01.1	PPB	12/7/2023 19:35	SMPL
O5494-19	MC0AE7	124	0.0559 -		SFAM01.1	PPB	12/7/2023 19:37	SMPL
CCV067	CCV067	24023	5.4223 -		SFAM01.1	PPB	12/7/2023 19:42	SMPL
CCB067	CCB067	-282	-0.0353 -		SFAM01.1	PPB	12/7/2023 19:44	SMPL
O5494-20	MC0AE8	0	0.028 -		SFAM01.1	PPB	12/7/2023 19:47	SMPL
O5494-21	MC0AF0	17	0.0318 -		SFAM01.1	PPB	12/7/2023 19:49	SMPL
O5494-22	MC0AF2	40	0.037 -		SFAM01.1	PPB	12/7/2023 19:51	SMPL
PB157632BL	PBW632	22	0.0329 -		SFAM01.1	PPB	12/7/2023 19:54	SMPL
O5495-01	MC0AL9	29	0.0345 -		SFAM01.1	PPB	12/7/2023 19:56	SMPL
O5495-02	MC0AM0	108	0.0523 -		SFAM01.1	PPB	12/7/2023 19:58	SMPL
O5495-03	MC0AM0D	118	0.0545 -		SFAM01.1	PPB	12/7/2023 20:00	SMPL
O5495-04	MC0AM0S	3636	0.8445 -		SFAM01.1	PPB	12/7/2023 20:03	SMPL
O5496-01	MC0AK5	-22	0.0231 -		SFAM01.1	PPB	12/7/2023 20:05	SMPL
O5496-02	MC0AK6	24	0.0334 -		SFAM01.1	PPB	12/7/2023 20:07	SMPL
CCV068	CCV068	20151	4.5528 -		SFAM01.1	PPB	12/7/2023 20:09	SMPL
CCB068	CCB068	-202	-0.0173 -		SFAM01.1	PPB	12/7/2023 20:12	SMPL