

7470A


LB128628

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41920



A= 0.0000e+000

B= 2.3793e-004

C= -4.2533e-002

Rho= 0.9999066

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.031	-0.031	49	0.000	49	0				
0.2	0.200	0.210	0.010	1063	0.0 %	1063					
2.5	2.500	2.507	0.007	10716	0.0 %	10716					
5.0	5.000	4.986	-0.014	21133	0.0 %	21133					
7.5	7.500	7.596	0.096	32105	0.0 %	32105					
10.0	10.000	9.931	-0.069	41920	0.0 %	41920					

100057
100057

LB128628 INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	49	-		0 7470A	PPB	12/08/2023 09:42	Std
	0.2 S0.2	1063	-		0.2 7470A	PPB	12/08/2023 09:45	Std
	2.5 S2.5	10716	-		2.5 7470A	PPB	12/08/2023 09:47	Std
	5 S5.0	21133	-		5 7470A	PPB	12/08/2023 09:49	Std
	7.5 S7.5	32105	-		7.5 7470A	PPB	12/08/2023 09:51	Std
	10 S10.0	41920	-		10 7470A	PPB	12/08/2023 10:02	Std
ICV12	ICV12	15626	3.6754	-	7470A	PPB	12/08/2023 10:05	SMPL
ICB12	ICB12	-7	-0.0442	-	7470A	PPB	12/08/2023 10:10	SMPL
CCV69	CCV69	20707	4.8843	-	7470A	PPB	12/08/2023 10:15	SMPL
CCB69	CCB69	12	-0.0397	-	7470A	PPB	12/08/2023 10:20	SMPL
CRA	CRA	1042	0.2054	-	7470A	PPB	12/08/2023 10:27	SMPL
HighStd	HighStd	38489	9.1152	-	7470A	PPB	12/08/2023 10:38	SMPL
ChkStd	ChkStd	33582	7.9476	-	7470A	PPB	12/08/2023 10:46	SMPL
PB157615BL	PBW	19	-0.038	-	7470A	PPB	12/08/2023 10:51	SMPL
PB157615BS	LCSW	17315	4.0772	-	7470A	PPB	12/08/2023 10:54	SMPL
O5599-02	LKWD-BACKFILL-01	-190	-0.0877	-	7470A	PPB	12/08/2023 10:56	SMPL
O5599-04	LKWD-BACKFILL-02	928	0.1783	-	7470A	PPB	12/08/2023 10:58	SMPL
O5599-06	LKWD-BACKFILL-03	3169	0.7115	-	7470A	PPB	12/08/2023 11:01	SMPL
O5599-08	LKWD-BACKFILL-04	-82	-0.062	-	7470A	PPB	12/08/2023 11:03	SMPL
O5599-10	LKWD-BACKFILL-05	72	-0.0254	-	7470A	PPB	12/08/2023 11:05	SMPL
CCV70	CCV70	21936	5.1767	-	7470A	PPB	12/08/2023 11:20	SMPL
CCB70	CCB70	-93	-0.0647	-	7470A	PPB	12/08/2023 11:22	SMPL
O5599-12	LKWD-BACKFILL-06	92	-0.0206	-	7470A	PPB	12/08/2023 11:25	SMPL
O5599-14	LKWD-BACKFILL-07	113	-0.0156	-	7470A	PPB	12/08/2023 11:27	SMPL
O5599-16	LKWD-BACKFILL-08	464	0.0679	-	7470A	PPB	12/08/2023 11:29	SMPL
O5599-16DUP	LKWD-BACKFILL-08DUP	305	0.03	-	7470A	PPB	12/08/2023 11:31	SMPL
O5599-16MS	LKWD-BACKFILL-08MS	18698	4.4063	-	7470A	PPB	12/08/2023 11:34	SMPL
O5599-16MSD	LKWD-BACKFILL-08MSD	19033	4.486	-	7470A	PPB	12/08/2023 11:36	SMPL
O5600-02	TOPSOIL-LKWD-DENALI-1	328	0.0355	-	7470A	PPB	12/08/2023 11:38	SMPL
O5600-04	TOPSOIL-LKWD-DENALI-2	366	0.0445	-	7470A	PPB	12/08/2023 11:40	SMPL
O5600-06	TOPSOIL-LKWD-DENALI-3	365	0.0443	-	7470A	PPB	12/08/2023 11:43	SMPL
PB157616BL	PBW	102	-0.0183	-	7470A	PPB	12/08/2023 11:45	SMPL
CCV71	CCV71	19574	4.6147	-	7470A	PPB	12/08/2023 11:51	SMPL
CCB71	CCB71	-160	-0.0806	-	7470A	PPB	12/08/2023 11:53	SMPL
PB157616BS	LCSW	17647	4.1562	-	7470A	PPB	12/08/2023 11:58	SMPL
O5628-02	SB-01-COMP	143	-0.0085	-	7470A	PPB	12/08/2023 12:00	SMPL
O5628-02DUP	SB-01-COMPDUP	49	-0.0309	-	7470A	PPB	12/08/2023 12:02	SMPL
O5628-02MS	SB-01-COMPMS	15450	3.6335	-	7470A	PPB	12/08/2023 12:05	SMPL
O5628-02MSD	SB-01-COMPMSD	14377	3.3782	-	7470A	PPB	12/08/2023 12:07	SMPL
O5628-04	SB-02-COMP	-281	-0.1094	-	7470A	PPB	12/08/2023 12:09	SMPL
O5628-06	SB-04-COMP	-56	-0.0559	-	7470A	PPB	12/08/2023 12:12	SMPL
O5628-08	SB-03-COMP	-26	-0.0487	-	7470A	PPB	12/08/2023 12:14	SMPL
O5629-03	VNJ-216	22	-0.0373	-	7470A	PPB	12/08/2023 12:16	SMPL
O5629-06	72-11985	-20	-0.0473	-	7470A	PPB	12/08/2023 12:18	SMPL
CCV72	CCV72	22408	5.289	-	7470A	PPB	12/08/2023 12:30	SMPL
CCB72	CCB72	-262	-0.1049	-	7470A	PPB	12/08/2023 12:33	SMPL
O5635-05	M00-23-00024	352	0.0412	-	7470A	PPB	12/08/2023 12:35	SMPL
O5635-06	COMP-1	655	0.1133	-	7470A	PPB	12/08/2023 12:37	SMPL
O5635-07	COMP-2	381	0.0481	-	7470A	PPB	12/08/2023 12:39	SMPL
PB157618BL	PBW	340	0.0384	-	7470A	PPB	12/08/2023 12:42	SMPL
PB157618BS	LCSW	19863	4.6835	-	7470A	PPB	12/08/2023 12:44	SMPL
O5639-02	E8965C-CC-001	8050	1.8728	-	7470A	PPB	12/08/2023 12:46	SMPL
O5657-02	205-199	8	-0.0406	-	7470A	PPB	12/08/2023 12:48	SMPL
O5671-04	TP-34	96	-0.0197	-	7470A	PPB	12/08/2023 12:51	SMPL
O5671-08	TP-35	237	0.0139	-	7470A	PPB	12/08/2023 12:53	SMPL
O5671-08DUP	TP-35DUP	226	0.0112	-	7470A	PPB	12/08/2023 12:55	SMPL
CCV73	CCV73	19958	4.7061	-	7470A	PPB	12/08/2023 12:58	SMPL
CCB73	CCB73	-156	-0.0796	-	7470A	PPB	12/08/2023 13:00	SMPL
O5671-08MS	TP-35MS	17619	4.1496	-	7470A	PPB	12/08/2023 13:02	SMPL
O5671-08MSD	TP-35MSD	17924	4.2221	-	7470A	PPB	12/08/2023 13:07	SMPL
PB157615TB	PB157615TB	-191	-0.088	-	7470A	PPB	12/08/2023 13:09	SMPL
PB157576TB	PB157576TB	333	0.0367	-	7470A	PPB	12/08/2023 13:12	SMPL
PB157592TB	PB157592TB	402	0.0531	-	7470A	PPB	12/08/2023 13:14	SMPL
CCV74	CCV74	21658	5.1106	-	7470A	PPB	12/08/2023 13:16	SMPL
CCB74	CCB74	-83	-0.0623	-	7470A	PPB	12/08/2023 13:19	SMPL
O5599-16LX5		-33	-0.0504	-	7470A	PPB	12/08/2023 13:21	SMPL
O5599-16A		17986	4.2369	-	7470A	PPB	12/08/2023 13:23	SMPL
O5628-02LX5		-152	-0.0787	-	7470A	PPB	12/08/2023 13:25	SMPL
O5628-02A		18806	4.432	-	7470A	PPB	12/08/2023 13:28	SMPL
O5671-08LX5		-149	-0.078	-	7470A	PPB	12/08/2023 13:30	SMPL
O5671-08A		17993	4.2385	-	7470A	PPB	12/08/2023 13:32	SMPL
CCV75	CCV75	21859	5.1584	-	7470A	PPB	12/08/2023 13:35	SMPL
CCB75	CCB75	-213	-0.0932	-	7470A	PPB	12/08/2023 13:37	SMPL