



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB128632

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	12/08/2023	13:00
2	CAL2	1	Water	NA	NA	20.5	7.00	12/08/2023	13:05
3	CAL3	1	Water	NA	NA	20.5	10.02	12/08/2023	13:06
4	ICV	1	Water	NA	NA	20.5	7.00	12/08/2023	13:10
5	CCV1	1	Water	NA	NA	20.6	2.01	12/08/2023	13:15
6	O5717-02	1	Water	NA	NA	20.8	6.18	12/08/2023	13:25
7	O5717-03	1	Water	NA	NA	20.6	6.39	12/08/2023	13:30
8	O5724-01	1	Water	NA	NA	20.2	6.88	12/08/2023	13:35
9	O5724-02	1	Water	NA	NA	20.7	5.87	12/08/2023	13:40
10	O5727-01	1	Water	NA	NA	20.1	5.87	12/08/2023	13:50
11	O5727-02	1	Water	NA	NA	20.4	5.86	12/08/2023	14:00
12	O5727-02DUP	1	Water	NA	NA	20.5	5.87	12/08/2023	14:02
13	CCV2	1	Water	NA	NA	20.5	12.02	12/08/2023	14:05

28128632

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O5724 WorkList ID : 176171 Department : Wet-Chemistry Date : 12-08-2023 12:35:27

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5717-02	TT-058-IDWGW-20231207	Water	pH	Cool 4 deg C	TETRO6	L11	12/07/2023	9040C
O5717-03	TT-59-IDWGW-20231207	Water	pH	Cool 4 deg C	TETRO6	L11	12/07/2023	9040C
O5724-01	MINI-CELL	Water	pH	Cool 4 deg C	STER14	L11	12/04/2023	9040C
O5724-02	JS10000	Water	pH	Cool 4 deg C	STER14	L11	12/04/2023	9040C
O5727-01	POOL-WATER	Water	pH	Cool 4 deg C	STER06	L11	12/05/2023	9040C
O5727-02	POOL-WATER	Water	pH	Cool 4 deg C	STER06	L11	12/05/2023	9040C

Date/Time 12/08/23 12:40
Raw Sample Received by: 28128632
Raw Sample Relinquished by: RM San

Date/Time 12/08/23 15:00
Raw Sample Received by: RM San
Raw Sample Relinquished by: 28128632