



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB128634

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.00	12/08/2023	14:15
2	CAL2	1	Water	NA	NA	20.4	7.01	12/08/2023	14:16
3	CAL3	1	Water	NA	NA	20.3	10.02	12/08/2023	14:20
4	ICV	1	Water	NA	NA	20.4	6.98	12/08/2023	14:25
5	CCV1	1	Water	NA	NA	20.5	2.01	12/08/2023	14:30
6	O5718-04	1	Water	NA	NA	20.1	6.67	12/08/2023	14:35
7	O5718-04DUP	1	Water	NA	NA	20.2	6.68	12/08/2023	14:36
8	CCV2	1	Water	NA	NA	20.5	12.02	12/08/2023	14:40

WORKLIST(Hardcopy Internal Chain)

VB 128634

WorkList Name : corosovity O5718 water **WorkList ID :** 176177 **Department :** Wet-Chemistry **Date :** 12-08-2023 13:58:20

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5718-04	WC-LIQUID-20231206	Water	Corrosivity	Cool 4 deg C	PARS02	L21	12/06/2023	9040C

Date/Time 12/08/23 14:10
Raw Sample Received by: JH GRC / RM SM
Raw Sample Relinquished by: RM SM

Date/Time 12/08/23 16:00
Raw Sample Received by: RM SM
Raw Sample Relinquished by: JH GRC