



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128637

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	12/08/2023	09:20
2	O5635-08		1	12	100.00	>212.0	>212.0	12/08/2023	09:50
3	O5635-08DUP		1	12	100.00	>212.0	>212.0	12/08/2023	10:20
4	O5650-01		1	14	100.00	>212.0	>212.0	12/08/2023	10:50
5	O5650-02		1	12	100.00	>212.0	>212.0	12/08/2023	11:20
6	O5650-03		1	13	100.00	>212.0	>212.0	12/08/2023	11:50
7	O5650-04		1	14	100.00	>212.0	>212.0	12/08/2023	12:20
8	O5658-02		1	12	100.00	>212.0	>212.0	12/08/2023	12:50
9	O5658-03		1	12	100.00	>212.0	>212.0	12/08/2023	13:20
10	O5660-01		1	14	100.00	>212.0	>212.0	12/08/2023	13:50
11	O5660-02		1	12	100.00	>212.0	>212.0	12/08/2023	14:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6128637

WorkList Name : FP 12-06 WorkList ID : 176103 Department : Wet-Chemistry Date : 12-06-2023 17:03:04

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5635-08C	COMP-3	Water	Flash Point	Cool 4 deg C	PSEG03	L31	12/05/2023	1010B
O5650-01D	50194	Water	Flash Point	Cool 4 deg C	PSEG03	L31	12/06/2023	1010B
O5650-02D	50195	Water	Flash Point	Cool 4 deg C	PSEG03	L31	12/06/2023	1010B
O5650-03D	50196	Water	Flash Point	Cool 4 deg C	PSEG03	L31	12/06/2023	1010B
O5650-04F	50197	Water	Flash Point	Cool 4 deg C	PSEG03	L31	12/06/2023	1010B
O5658-02C	34147	Water	Flash Point	Cool 4 deg C	PSEG03	L21	12/06/2023	1010B
O5658-03C	60118	Water	Flash Point	Cool 4 deg C	PSEG03	L21	12/06/2023	1010B
O5660-01A	1705	Water	Flash Point	Cool 4 deg C	PSEG03	L21	12/06/2023	1010B
O5660-02C	1100	Water	Flash Point	Cool 4 deg C	PSEG03	L21	12/06/2023	1010B

Date/Time 12/08/2023 09:10
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM CWC

Date/Time 12/08/2023 14:30
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM CWC