

LB128638


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

50573



A= 0.0000e+000

B= 1.9980e-004 slope

C= -9.9585e-002 y-intercept

Rho= 0.9996618

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/1
0.00	0.000	0.041	0.041	702	0.000	702					
0.05	0.050					442	0				
0.20	0.200	0.228	0.028	1638	0.0 %	1638					
2.50	2.500	2.337	-0.163	12195	0.0 %	12195					
5.00	5.000	5.150	0.150	26275	0.0 %	26275					
7.50	7.500	7.439	-0.061	37732	0.0 %	37732					
10.0	10.000	10.005	0.005	50573	0.0 %	50573					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	702	-		0 SFAM01.1	PPB	12/8/2023 16:14	Std
	0.2 S01	1638	-		0.2 SFAM01.1	PPB	12/8/2023 16:16	Std
	2.5 S02	12195	-		2.5 SFAM01.1	PPB	12/8/2023 16:21	Std
	5 S03	26275	-		5 SFAM01.1	PPB	12/8/2023 16:23	Std
	7.5 S04	37732	-		7.5 SFAM01.1	PPB	12/8/2023 16:25	Std
	10 S05	50573	-		10 SFAM01.1	PPB	12/8/2023 16:27	Std
ICV013	ICV013	20288	3.954 -		SFAM01.1	PPB	12/8/2023 16:31	SMPL
ICB013	ICB013	401	-0.0195 -		SFAM01.1	PPB	12/8/2023 16:33	SMPL
CCV076	CCV076	26391	5.1734 -		SFAM01.1	PPB	12/8/2023 16:35	SMPL
CCB076	CCB076	389	-0.0219 -		SFAM01.1	PPB	12/8/2023 16:37	SMPL
PB157649BL	PBW649	347	-0.0303 -		SFAM01.1	PPB	12/8/2023 16:40	SMPL
O5528-01	MC0AJ2	1620	0.2241 -		SFAM01.1	PPB	12/8/2023 16:42	SMPL
O5528-02	MC0AJ2D	1753	0.2507 -		SFAM01.1	PPB	12/8/2023 16:44	SMPL
O5528-03	MC0AJ2S	6357	1.1706 -		SFAM01.1	PPB	12/8/2023 16:46	SMPL
O5530-01	MC0AJ4	833	0.0669 -		SFAM01.1	PPB	12/8/2023 16:49	SMPL
O5530-02	MC0AJ4D	931	0.0864 -		SFAM01.1	PPB	12/8/2023 16:51	SMPL
O5530-03	MC0AJ4S	6489	1.1969 -		SFAM01.1	PPB	12/8/2023 16:53	SMPL
O5583-01	MC0AJ6	785	0.0573 -		SFAM01.1	PPB	12/8/2023 16:56	SMPL
O5583-02	MC0AJ6D	797	0.0597 -		SFAM01.1	PPB	12/8/2023 16:58	SMPL
O5583-03	MC0AJ6S	5904	1.08 -		SFAM01.1	PPB	12/8/2023 17:00	SMPL
O5584-01	MC0AJ7	822	0.0647 -		SFAM01.1	PPB	12/8/2023 17:03	SMPL
O5584-02	MC0AJ7D	832	0.0667 -		SFAM01.1	PPB	12/8/2023 17:08	SMPL
O5584-03	MC0AJ7S	6035	1.1062 -		SFAM01.1	PPB	12/8/2023 17:10	SMPL
CCV077	CCV077	26686	5.2323 -		SFAM01.1	PPB	12/8/2023 17:12	SMPL
CCB077	CCB077	548	0.0099 -		SFAM01.1	PPB	12/8/2023 17:14	SMPL