



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB128642

Slope : 99.2

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	12/08/2023	14:30
2	CAL2	1	Water	NA	NA	20.5	6.99	12/08/2023	14:31
3	CAL3	1	Water	NA	NA	20.4	10.02	12/08/2023	14:33
4	ICV	1	Water	NA	NA	20.5	7.01	12/08/2023	14:35
5	CCV1	1	Water	NA	NA	20.5	2.01	12/08/2023	14:40
6	O5718-02	1	Solid	20.02	20	23.1	6.75	12/08/2023	14:45
7	O5718-02DUP	1	Solid	20.03	20	23.2	6.76	12/08/2023	14:46
8	O5726-04	1	Solid	20.02	20	23.8	7.49	12/08/2023	14:50
9	O5726-08	1	Solid	20.03	20	23.6	7.44	12/08/2023	14:55
10	CCV2	1	Water	NA	NA	20.5	12.02	12/08/2023	15:00

WORKLIST(Hardcopy Internal Chain)

128642

WorkList Name : corosovity O5726 WorkList ID : 176176 Department : Wet-Chemistry Date : 12-08-2023 13:57:37

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5718-02	WC-SOIL-20231206	Solid	Corrosivity	Cool 4 deg C	PARS02	L21	12/06/2023	9045D
O5726-04	WC-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	L21	12/08/2023	9045D
O5726-08	WC-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	L21	12/08/2023	9045D

Date/Time 12/08/23 14:20
 Raw Sample Received by: JD (WC)
 Raw Sample Relinquished by: JD (WC)

Date/Time 12/08/23 17:30
 Raw Sample Received by: JD (WC)
 Raw Sample Relinquished by: JD (WC)