


LB128643

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.0340e-004 slope

C= -7.4251e-002 y-intercept

Rho= 0.9996820

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%1)
0.00	0.000	0.052	0.052	622	0.000	622					=
0.05	0.050					442					27
0.20	0.200	0.253	0.053	1607	0.0 %	1607	0				-5
2.50	2.500	2.366	-0.134	11997	0.0 %	11997					1
5.00	5.000	5.067	0.067	25277	0.0 %	25277					-2
7.50	7.500	7.373	-0.127	36614	0.0 %	36614					1
10.0	10.000	10.089	0.089	49965	0.0 %	49965					

LB128643

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	622	-		0 SFAM01.1	PPB	12/8/2023 18:07	Std
	0.2 S01	1607	-		0.2 SFAM01.1	PPB	12/8/2023 18:09	Std
	2.5 S02	11997	-		2.5 SFAM01.1	PPB	12/8/2023 18:11	Std
	5 S03	25277	-		5 SFAM01.1	PPB	12/8/2023 18:13	Std
	7.5 S04	36614	-		7.5 SFAM01.1	PPB	12/8/2023 18:16	Std
	10 S05	49965	-		10 SFAM01.1	PPB	12/8/2023 18:18	Std
ICV014	ICV014	20114	4.017 -		SFAM01.1	PPB	12/8/2023 18:21	SMPL
ICB014	ICB014	351	-0.0029 -		SFAM01.1	PPB	12/8/2023 18:23	SMPL
CCV078	CCV078	25445	5.1013 -		SFAM01.1	PPB	12/8/2023 18:25	SMPL
CCB078	CCB078	352	-0.0027 -		SFAM01.1	PPB	12/8/2023 18:28	SMPL
PB157665BL	PBW665	314	-0.0104 -		SFAM01.1	PPB	12/8/2023 18:30	SMPL
PB157446TB	LEB446	619	0.0517 -		SFAM01.1	PPB	12/8/2023 18:32	SMPL
O5511-01	MJLAE0	1213	0.1725 -		SFAM01.1	PPB	12/8/2023 18:40	SMPL
O5511-02	MJLAE0D	978	0.1247 -		SFAM01.1	PPB	12/8/2023 18:42	SMPL
O5511-03	MJLAE0S	6408	1.2292 -		SFAM01.1	PPB	12/8/2023 18:45	SMPL
CCV079	CCV079	25725	5.1583 -		SFAM01.1	PPB	12/8/2023 18:47	SMPL
CCB079	CCB079	525	0.0325 -		SFAM01.1	PPB	12/8/2023 18:49	SMPL