



PERCENT SOLID

Supervisor: Iwona  
Analyst: JIGNESH  
Date: 12/12/2023

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:00  
In Date: 12/11/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:14  
Out Date: 12/12/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB128660

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments        |
|----------|-----------------|--------|----------------|--------------|-------------------------|---------------------------|---------|-----------------|
| O5742-06 | 50176           | 1      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | OIL SAMPLE      |
| O5745-01 | WC-1            | 2      | 1.18           | 8.43         | 9.61                    | 8.23                      | 83.6    |                 |
| O5745-02 | WC-1-EPH        | 3      | 1.15           | 8.38         | 9.53                    | 8.07                      | 82.6    |                 |
| O5745-03 | WC-1-VOC        | 4      | 1.19           | 8.69         | 9.88                    | 8.08                      | 79.3    |                 |
| O5763-01 | 01-A-01-B-01-C  | 5      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5763-02 | 02-A-02-B-02-C  | 6      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5763-03 | 03-A-03-B-03-C  | 7      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5763-04 | 04-A-04-B-04-C  | 8      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5763-05 | 05-A-05-B-05-C  | 9      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5763-06 | 06-A-06-B-06-C  | 10     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5763-07 | 07-A-07-B-07-C  | 11     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5763-08 | 08-A-08-B-08-C  | 12     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5764-01 | PCB-120623-01   | 13     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5764-02 | PCB-120623-02   | 14     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |
| O5764-03 | PCB-120623-03   | 15     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CAULKING SAMPLE |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$