



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128662

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	12/11/2023	09:15
2	O5691-03		1	12	100.00	>212.0	>212.0	12/11/2023	09:45
3	O5691-03DUP		1	12	100.00	>212.0	>212.0	12/11/2023	10:15
4	O5700-01		1	14	100.00	>212.0	>212.0	12/11/2023	10:45
5	O5718-04		1	12	100.00	>212.0	>212.0	12/11/2023	11:15
6	O5742-02		1	13	100.00	>212.0	>212.0	12/11/2023	13:10
7	O5742-03		1	12	100.00	>212.0	>212.0	12/11/2023	13:40
8	O5742-04		1	12	100.00	>212.0	>212.0	12/11/2023	14:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16128662

WorkList Name : FP 12-11

WorkList ID : 176211

Department : Wet-Chemistry

Date : 12-11-2023 08:39:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5691-03	A3909	Water	Flash Point	Cool 4 deg C	PSEG03	L21	12/07/2023	1010B
O5700-01	257419	Water	Flash Point	Cool 4 deg C	PSEG03	L31	12/07/2023	1010B
O5718-04	WC-LIQUID-20231206	Water	Flash Point	Cool 4 deg C	PARS02	L21	12/06/2023	1010B
O5742-02	50206	Water	Flash Point	Cool 4 deg C	PSEG03	L21	12/08/2023	1010B
O5742-03	COMP-1	Water	Flash Point	Cool 4 deg C	PSEG03	L21	12/08/2023	1010B
O5742-04	COMP-2	Water	Flash Point	Cool 4 deg C	PSEG03	L21	12/08/2023	1010B

Date/Time 12/11/2023 09:10
 Raw Sample Received by: RM gwc
 Raw Sample Relinquished by: gwc

Date/Time 12/11/2023 16:45
 Raw Sample Received by: gwc
 Raw Sample Relinquished by: RM gwc