

LB128664


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

46280



A= 0.0000e+000

B= 2.1148e-004

C= 5.0987e-002

Rho= 0.9981171

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0D
0.0	0.000	-0.008	-0.008	-280	0.000	-280	0				15
0.2	0.200	0.230	0.030	845	0.0 %	845					5
2.5	2.500	2.628	0.128	12186	0.0 %	12186					-7
5.0	5.000	4.643	-0.357	21714	0.0 %	21714					5
7.5	7.500	7.869	0.369	36970	0.0 %	36970					-2
10.0	10.000	9.838	-0.162	46280	0.0 %	46280					

LB128664 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Dec 2023 14:07:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-280	-	0.0000	7470A	PPB	11 Dec 2023 14:48:16	S	Std
0.2 - 1	50.2	845	-	0.2000	7470A	PPB	11 Dec 2023 14:50:34	S	Std
2.5 - 1	52.5	12186	-	2.5000	7470A	PPB	11 Dec 2023 14:52:51	S	Std
5.0 - 1	55	21714	-	5.0000	7470A	PPB	11 Dec 2023 14:55:08	S	Std
7.5 - 1	57.5	36970	-	7.5000	7470A	PPB	11 Dec 2023 14:57:26	S	Std
10.0 - 1	511	46280	-	10.0000	7470A	PPB	11 Dec 2023 15:05:11	S	Std
ICV16 - 1	ICV16	18878	4.0433	-7470A	PPB	PPB	11 Dec 2023 15:10:33	U	SMPL
ICB16 - 1	ICB16	-146	0.0201	-7470A	PPB	PPB	11 Dec 2023 15:12:48	U	SMPL
CCV83 - 1	CCV83	22885	4.8907	-7470A	PPB	PPB	11 Dec 2023 15:15:05	U	SMPL
CCB83 - 1	CCB83	-165	0.0161	-7470A	PPB	PPB	11 Dec 2023 15:17:20	U	SMPL
CRA - 1	CRA	846	0.2299	-7470A	PPB	PPB	11 Dec 2023 15:19:37	U	SMPL
HighStd - 1	HighStd	46363	9.8557	-7470A	PPB	PPB	11 Dec 2023 15:21:52	U	SMPL
ChkStd - 1	ChkStd	33330	7.0995	-7470A	PPB	PPB	11 Dec 2023 15:24:08	U	SMPL
PB157708BL - 1	PBW	-419	-0.0376	-7470A	PPB	PPB	11 Dec 2023 15:26:27	U	SMPL
PB157708BS - 1	LCSW	19106	4.0915	-7470A	PPB	PPB	11 Dec 2023 15:28:46	U	SMPL
O5643-02 - 1	COMP	411	0.1379	-7470A	PPB	PPB	11 Dec 2023 15:31:02	U	SMPL
O5643-02MS - 1	COMPMS	14995	3.2221	-7470A	PPB	PPB	11 Dec 2023 15:35:39	U	SMPL
O5643-02MSD - 1	COMPMSD	13544	2.9152	-7470A	PPB	PPB	11 Dec 2023 15:37:58	U	SMPL
O5691-03 - 1	A3909	-278	-0.0078	-7470A	PPB	PPB	11 Dec 2023 15:40:17	U	SMPL
CCV84 - 1	CCV84	22236	4.7534	-7470A	PPB	PPB	11 Dec 2023 15:42:35	U	SMPL
CCB84 - 1	CCB84	-130	0.0235	-7470A	PPB	PPB	11 Dec 2023 15:44:51	U	SMPL
O5700-01 - 1	257419	1359	0.3384	-7470A	PPB	PPB	11 Dec 2023 15:47:09	U	SMPL
O5717-02 - 1	TT-058-IDWGW-20231207	222	0.0979	-7470A	PPB	PPB	11 Dec 2023 15:49:24	U	SMPL
O5717-03 - 1	TT-059-IDWGW-20231207	366	0.1284	-7470A	PPB	PPB	11 Dec 2023 15:51:39	U	SMPL
O5718-03 - 1	WC-LIQUID-20231206	1639	0.3976	-7470A	PPB	PPB	11 Dec 2023 15:53:55	U	SMPL
O5643-02DUP - 1	COMPDUP	582	0.1741	-7470A	PPB	PPB	11 Dec 2023 15:56:10	U	SMPL
O5643-02LX5 - 1		22	0.0556	-7470A	PPB	PPB	11 Dec 2023 15:58:27	U	SMPL
O5643-02A - 1		18208	3.9011	-7470A	PPB	PPB	11 Dec 2023 16:00:44	U	SMPL
CCV85 - 1	CCV85	23800	5.0842	-7470A	PPB	PPB	11 Dec 2023 16:03:01	U	SMPL
CCB85 - 1	CCB85	-80	0.0341	-7470A	PPB	PPB	11 Dec 2023 16:05:20	U	SMPL