



PERCENT SOLID

Supervisor: Iwona  
Analyst: JIGNESH  
Date: 12/12/2023

OVENTEMP IN Celsius(°C): 107  
Time IN: 15:25  
In Date: 12/11/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 07:55  
Out Date: 12/12/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB128666

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|-----------------|--------|----------------|--------------|-------------------------|---------------------------|---------|----------|
| O5746-01 | EZPQ0           | 1      | 1.18           | 8.59         | 9.77                    | 8.84                      | 89.2    |          |
| O5746-02 | EZPQ1           | 2      | 1.18           | 8.46         | 9.64                    | 8.16                      | 82.5    |          |
| O5746-03 | EZPQ2           | 3      | 1.15           | 8.83         | 9.98                    | 8.39                      | 82.0    |          |
| O5746-04 | EZPQ3           | 4      | 1.18           | 8.43         | 9.61                    | 7.99                      | 80.8    |          |
| O5746-05 | EZPQ4           | 5      | 1.19           | 8.79         | 9.98                    | 8.37                      | 81.7    |          |
| O5746-06 | EZPQ5           | 6      | 1.15           | 8.46         | 9.61                    | 8.09                      | 82.0    |          |
| O5746-07 | EZPQ6           | 7      | 1.18           | 8.56         | 9.74                    | 8.18                      | 81.8    |          |
| O5746-08 | EZPQ7           | 8      | 1.17           | 8.80         | 9.97                    | 8.56                      | 84.0    |          |
| O5746-09 | EZPQ8           | 9      | 1.19           | 8.49         | 9.68                    | 8.21                      | 82.7    |          |
| O5746-10 | EZPQ9           | 10     | 1.18           | 8.48         | 9.66                    | 8.22                      | 83.0    |          |
| O5746-11 | EZPR0           | 11     | 1.16           | 8.82         | 9.98                    | 8.48                      | 83.0    |          |
| O5746-12 | EZPR1           | 12     | 1.16           | 8.78         | 9.94                    | 8.09                      | 78.9    |          |
| O5746-13 | EZPR2           | 13     | 1.12           | 8.81         | 9.93                    | 8.23                      | 80.7    |          |
| O5746-14 | EZPR3           | 14     | 1.17           | 8.58         | 9.75                    | 7.99                      | 79.5    |          |
| O5746-15 | EZPR4           | 15     | 1.17           | 8.80         | 9.97                    | 8.42                      | 82.4    |          |
| O5746-16 | EZPR5           | 16     | 1.13           | 8.45         | 9.58                    | 8.43                      | 86.4    |          |
| O5746-17 | EZPR6           | 17     | 1.19           | 8.79         | 9.98                    | 8.21                      | 79.9    |          |
| O5746-18 | EZPR7           | 18     | 1.19           | 8.65         | 9.84                    | 8.38                      | 83.1    |          |
| O5746-19 | EZPR8           | 19     | 1.11           | 8.76         | 9.87                    | 8.16                      | 80.5    |          |
| O5746-20 | EZPS0           | 20     | 1.18           | 8.59         | 9.77                    | 8.06                      | 80.1    |          |
| O5746-21 | EZPS0D          | 21     | 1.18           | 8.59         | 9.77                    | 8.06                      | 80.1    |          |
| O5746-22 | EZPS0S          | 22     | 1.18           | 8.59         | 9.77                    | 8.06                      | 80.1    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$