



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128678

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/12/2023	08:00
2	CAL2	1	Water	NA	NA	20.3	7.01	12/12/2023	08:01
3	CAL3	1	Water	NA	NA	20.2	10.02	12/12/2023	08:02
4	ICV	1	Water	NA	NA	20.5	7.00	12/12/2023	08:05
5	CCV1	1	Water	NA	NA	20.3	2.01	12/12/2023	08:10
6	O5742-06	1	Solid	20.01	20	22.8	7.86	12/12/2023	08:20
7	O5742-06DUP	1	Solid	20.02	20	22.9	7.87	12/12/2023	08:21
8	CCV2	1	Water	NA	NA	20.6	12.02	12/12/2023	08:25

WORKLIST(Hardcopy Internal Chain)

VB 128678

WorkList Name : ph 05742 WorkList ID : 176245 Department : Wet-Chemistry Date : 12-12-2023 07:38:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5742-06	50176	Solid	pH	Cool 4 deg C	PSEG03	L21	12/08/2023	9045D

Date/Time 12-12-23 07:45
 Raw Sample Received by: JDCSM
 Raw Sample Relinquished by: JDCSM

Date/Time 12-12-23 11:00
 Raw Sample Received by: JDCSM
 Raw Sample Relinquished by: JDCSM