


L3128683

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.7113e-004

C= -1.5018e-002

Rho= 0.9994979

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% I)
0.0	0.000	-0.019	-0.019	-22	0.000	-22					
0.2	0.200	0.162	-0.038	1033	0.0 %	1033					-19
2.5	2.500	2.658	0.158	15621	0.0 %	15621					6
5.0	5.000	4.822	-0.178	28264	0.0 %	28264					-4
7.5	7.500	7.640	0.140	44730	0.0 %	44730					2
10.0	10.000	9.937	-0.063	58157	0.0 %	58157					-1

LB128683 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Dec 2023 11:18:48

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-22	-	0.0000	7471B	PPB	12 Dec 2023 11:26:47	S	Std
0.2 - 1	50.2	1033	-	0.2000	7471B	PPB	12 Dec 2023 11:29:05	S	Std
2.5 - 1	52.1	15621	-	2.5000	7471B	PPB	12 Dec 2023 11:35:52	S	Std
5.0 - 1	55	28264	-	5.0000	7471B	PPB	12 Dec 2023 11:38:14	S	Std
7.5 - 1	57.5	44730	-	7.5000	7471B	PPB	12 Dec 2023 11:40:33	S	Std
10.0 - 1	58.157	58157	-	10.0000	7471B	PPB	12 Dec 2023 11:43:00	S	Std
ICV17 - 1	ICV17	24132	4.1147	-	7471B	PPB	12 Dec 2023 11:48:35	U	SMPL
ICB17 - 1	ICB17	-340	-0.0732	-	7471B	PPB	12 Dec 2023 11:50:52	U	SMPL
CCV86 - 1	CCV86	28595	4.8784	-	7471B	PPB	12 Dec 2023 11:53:17	U	SMPL
CCB86 - 1	CCB86	-347	-0.0744	-	7471B	PPB	12 Dec 2023 11:55:32	U	SMPL
CRA - 1	CRA	1051	0.1648	-	7471B	PPB	12 Dec 2023 11:57:56	U	SMPL
HighStd - 1	HighStd	57881	9.8902	-	7471B	PPB	12 Dec 2023 12:00:11	U	SMPL
ChkStd - 1	ChkStd	41241	7.0426	-	7471B	PPB	12 Dec 2023 12:02:27	U	SMPL
PB157719BL - 1	PBS	-398	-0.0831	-	7471B	PPB	12 Dec 2023 12:04:58	U	SMPL
PB157719BS - 1	LCSS	24762	4.2225	-	7471B	PPB	12 Dec 2023 12:07:23	U	SMPL
O5718-01 - 1	WC-SOIL-20231206	8316	1.4081	-	7471B	PPB	12 Dec 2023 12:14:41	U	SMPL
O5718-01DUP - 1	WC-SOIL-20231206DUP	8283	1.4025	-	7471B	PPB	12 Dec 2023 12:16:57	U	SMPL
O5718-01MS - 1	WC-SOIL-20231206MS	38201	6.5223	-	7471B	PPB	12 Dec 2023 12:19:17	U	SMPL
O5718-01MSD - 1	WC-SOIL-20231206MSD	39655	6.7711	-	7471B	PPB	12 Dec 2023 12:21:37	U	SMPL
O5720-02 - 1	TT-FRACSEDIMENT-IDWSO-20231207	581	0.0844	-	7471B	PPB	12 Dec 2023 12:24:03	U	SMPL
CCV87 - 1	CCV87	29046	4.9556	-	7471B	PPB	12 Dec 2023 12:26:28	U	SMPL
CCB87 - 1	CCB87	-206	-0.0503	-	7471B	PPB	12 Dec 2023 12:28:44	U	SMPL
O5726-01 - 1	WC-1	9155	1.5517	-	7471B	PPB	12 Dec 2023 12:31:08	U	SMPL
O5726-05 - 1	WC-2	9527	1.6153	-	7471B	PPB	12 Dec 2023 12:33:23	U	SMPL
O5737-01 - 1	SU-02-12082023	1264	0.2013	-	7471B	PPB	12 Dec 2023 12:35:41	U	SMPL
O5740-01 - 1	CHRT-24328	4856	0.8160	-	7471B	PPB	12 Dec 2023 12:38:00	U	SMPL
O5740-03 - 1	RBR-251686	22279	3.7976	-	7471B	PPB	12 Dec 2023 12:40:16	U	SMPL
O5742-06 - 1	50176	96	0.0014	-	7471B	PPB	12 Dec 2023 12:42:34	U	SMPL
O5745-01 - 1	WC-1	5925	0.9989	-	7471B	PPB	12 Dec 2023 12:44:56	U	SMPL
O5718-01LX5 - 1		1734	0.2817	-	7471B	PPB	12 Dec 2023 12:47:13	U	SMPL
O5718-01A - 1		32159	5.4884	-	7471B	PPB	12 Dec 2023 12:49:32	U	SMPL
CCV88 - 1	CCV88	29593	5.0492	-	7471B	PPB	12 Dec 2023 12:51:48	U	SMPL
CCB88 - 1	CCB88	-166	-0.0434	-	7471B	PPB	12 Dec 2023 12:54:11	U	SMPL