



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB128689

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	12/12/2023	14:10
2	CAL2	1	Water	NA	NA	20.5	7.00	12/12/2023	14:11
3	CAL3	1	Water	NA	NA	20.5	10.02	12/12/2023	14:15
4	ICV	1	Water	NA	NA	20.5	7.01	12/12/2023	14:20
5	CCV1	1	Water	NA	NA	20.5	2.01	12/12/2023	14:25
6	O5768-04	1	Water	NA	NA	23.7	6.93	12/12/2023	14:30
7	O5768-04DUP	1	Water	NA	NA	23.8	6.94	12/12/2023	14:31
8	O5772-01	1	Water	NA	NA	22.8	6.94	12/12/2023	14:35
9	O5772-02	1	Water	NA	NA	23.7	6.72	12/12/2023	14:40
10	CCV2	1	Water	NA	NA	20.5	12.02	12/12/2023	14:45

NB 128689

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O5772

WorkList ID : 176286

Department : Wet-Chemistry

Date : 12-12-2023 13:45:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5768-04	ELZ-23-00112	Water	pH	Cool 4 deg C	PSEG03	L31	12/11/2023	9040C
O5772-01	OMEGA CELL SAMPLE	Water	pH	Cool 4 deg C	STER15	L11	12/08/2023	9040C
O5772-02	OMEGA CELL SAMPLE	Water	pH	Cool 4 deg C	STER15	L11	12/08/2023	9040C

Date/Time

12.12.23 13:55

Raw Sample Received by:

JD (WCS)

Raw Sample Relinquished by:

RM SM

Date/Time

12-12-23

Raw Sample Received by:

RM SM

Raw Sample Relinquished by:

JD (WCS)