



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB128693

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	12/12/2023	14:10
2	CAL2	1	Water	NA	NA	20.5	7.00	12/12/2023	14:11
3	CAL3	1	Water	NA	NA	20.5	10.02	12/12/2023	14:15
4	ICV	1	Water	NA	NA	20.5	7.01	12/12/2023	14:20
5	CCV1	1	Water	NA	NA	20.4	2.01	12/12/2023	14:50
6	O5767-02	1	Solid	20.02	20	22.6	8.41	12/12/2023	14:51
7	O5767-02DUP	1	Solid	20.03	20	22.7	8.42	12/12/2023	14:52
8	O5767-05	1	Solid	20.04	20	22.9	8.69	12/12/2023	15:00
9	O5767-08	1	Solid	20.02	20	22.8	8.73	12/12/2023	15:05
10	O5768-01	1	Solid	20.03	20	23.7	7.49	12/12/2023	15:10
11	O5768-02	1	Solid	20.02	20	22.2	7.70	12/12/2023	15:15
12	O5768-03	1	Solid	20.03	20	22.4	7.83	12/12/2023	15:20
13	CCV2	1	Water	NA	NA	20.6	12.02	12/12/2023	15:25

WORKLIST(Hardcopy Internal Chain)

128693

WorkList Name : corosovity O5758 WorkList ID : 176283 Department : Wet-Chemistry Date : 12-12-2023 13:39:13

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5767-02	SP-E1-COMP	Solid	Corrosivity	Cool 4 deg C	EART12	L11	12/11/2023	9045D
O5767-05	SP-E2-COMP	Solid	Corrosivity	Cool 4 deg C	EART12	L11	12/11/2023	9045D
O5767-08	SP-E3-COMP	Solid	Corrosivity	Cool 4 deg C	EART12	L11	12/11/2023	9045D
O5768-01	ELZ-COMP	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	12/11/2023	9045D
O5768-02	ELZ-23-00118	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	12/11/2023	9045D
O5768-03	ELZ-23-00109	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	12/11/2023	9045D

Date/Time 12-12-23 13:50
 Raw Sample Received by: JH (WCH)
 Raw Sample Relinquished by: RM SM

Date/Time 12-12-23 18:00
 Raw Sample Received by: RM SM
 Raw Sample Relinquished by: JH (WCH)