

LB128697


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

48671



A= 0.0000e+000

B= 2.0624e-004 slope

C= -6.0117e-003 *+10% intercept*

Rho= 0.9999488

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.004	-0.004	10	0.000	10					
0.05	0.050					442	0				
0.20	0.200	0.206	0.006	1029	0.0 %	1029					
2.50	2.500	2.492	-0.008	12111	0.0 %	12111					
5.00	5.000	5.046	0.046	24493	0.0 %	24493					
7.50	7.500	7.428	-0.072	36046	0.0 %	36046					
10.0	10.000	10.032	0.032	48671	0.0 %	48671					

0.1-0.311

LB128697

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	10	-		0 SFAM01.1	PPB	12/12/2023 16:23	Std
	0.2 S01	1029	-		0.2 SFAM01.1	PPB	12/12/2023 16:25	Std
	2.5 S02	12111	-		2.5 SFAM01.1	PPB	12/12/2023 16:27	Std
	5 S03	24493	-		5 SFAM01.1	PPB	12/12/2023 16:29	Std
	7.5 S04	36046	-		7.5 SFAM01.1	PPB	12/12/2023 16:32	Std
	10 S05	48671	-		10 SFAM01.1	PPB	12/12/2023 16:34	Std
ICV018	ICV018	18650	3.8404 -		SFAM01.1	PPB	12/12/2023 16:37	SMPL
ICB018	ICB018	-248	-0.0572 -		SFAM01.1	PPB	12/12/2023 16:39	SMPL
CCV089	CCV089	24593	5.0662 -		SFAM01.1	PPB	12/12/2023 16:41	SMPL
CCB089	CCB089	-226	-0.0526 -		SFAM01.1	PPB	12/12/2023 16:44	SMPL
PB157740BL	PBW740	-17	-0.0095 -		SFAM01.1	PPB	12/12/2023 16:46	SMPL
O4696-01	MDL-WATER	300	0.0559 -		SFAM01.1	PPB	12/12/2023 16:48	SMPL
CCV090	CCV090	24651	5.0781 -		SFAM01.1	PPB	12/12/2023 16:50	SMPL
CCB090	CCB090	-218	-0.051 -		SFAM01.1	PPB	12/12/2023 16:53	SMPL