



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB128698

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	12/12/2023	10:20
2	O5742-06		1	12	100.00	>212.0	>212.0	12/12/2023	10:50
3	O5742-06DUP		1	12	100.00	>212.0	>212.0	12/12/2023	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6128698

WorkList Name : fp s-12-12 WorkList ID : 176255 Department : Wet-Chemistry Date : 12-12-2023 08:50:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5742-06	50176	Solid	Flash Point	Cool 4 deg C	PSEG03	L21	12/08/2023	1010B

Date/Time 12/12/2023 10:10
 Raw Sample Received by: RAY CUNY
 Raw Sample Relinquished by: JH GOC

Date/Time 12/12/2023 12:00
 Raw Sample Received by: JH CUL
 Raw Sample Relinquished by: RAY CUNY