


LB128700

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.7665e-004 *slope*C= 5.0750e-003 *y-intercept*

Rho= 0.9999135

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>%D</i>
0.00	0.000	0.025	0.025	110	0.000	110					5
0.05	0.050					442	0				10
0.20	0.200	0.220	0.020	1219	0.0 %	1219					6
2.50	2.500	2.507	0.007	14162	0.0 %	14162					1
5.00	5.000	4.932	-0.068	27889	0.0 %	27889					1
7.50	7.500	7.444	-0.056	42112	0.0 %	42112					1
10.0	10.000	10.073	0.073	56992	0.0 %	56992					1

LB128700

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	110	-		0 SFAM01.1	PPB	12/12/2023 17:52	Std
	0.2 S01	1219	-		0.2 SFAM01.1	PPB	12/12/2023 17:54	Std
	2.5 S02	14162	-		2.5 SFAM01.1	PPB	12/12/2023 17:58	Std
	5 S03	27889	-		5 SFAM01.1	PPB	12/12/2023 18:01	Std
	7.5 S04	42112	-		7.5 SFAM01.1	PPB	12/12/2023 18:03	Std
	10 S05	56992	-		10 SFAM01.1	PPB	12/12/2023 18:05	Std
ICV019	ICV019	21475	3.7986 -		SFAM01.1	PPB	12/12/2023 18:09	SMPL
ICB019	ICB019	-185	-0.0276 -		SFAM01.1	PPB	12/12/2023 18:11	SMPL
CCV091	CCV091	28141	4.9761 -		SFAM01.1	PPB	12/12/2023 18:13	SMPL
CCB091	CCB091	-221	-0.034 -		SFAM01.1	PPB	12/12/2023 18:15	SMPL
PB157743BL	PBS743	39	0.012 -		SFAM01.1	PPB	12/12/2023 18:18	SMPL
O5680-01	E0XA7	2686	0.4796 -		SFAM01.1	PPB	12/12/2023 18:20	SMPL
O5680-02	E0XA8	977	0.1777 -		SFAM01.1	PPB	12/12/2023 18:22	SMPL
O5680-03	E0XA9	916	0.1669 -		SFAM01.1	PPB	12/12/2023 18:25	SMPL
O5680-04	E0XB0	304	0.0588 -		SFAM01.1	PPB	12/12/2023 18:27	SMPL
O5680-05	E0XM1	147	0.031 -		SFAM01.1	PPB	12/12/2023 18:29	SMPL
O5680-06	E0XB1	891	0.1625 -		SFAM01.1	PPB	12/12/2023 18:31	SMPL
O5680-07	E0XB2	445	0.0837 -		SFAM01.1	PPB	12/12/2023 18:34	SMPL
O5680-08	E0XL0	184	0.0376 -		SFAM01.1	PPB	12/12/2023 18:36	SMPL
O5680-09	E0XL1	469	0.0879 -		SFAM01.1	PPB	12/12/2023 18:38	SMPL
O5680-10	E0XL2	104	0.0234 -		SFAM01.1	PPB	12/12/2023 18:40	SMPL
O5680-11	E0XM2	598	0.1107 -		SFAM01.1	PPB	12/12/2023 18:43	SMPL
O5680-12	E0XS5	578	0.1072 -		SFAM01.1	PPB	12/12/2023 18:45	SMPL
O5680-13	E0XS6	104	0.0234 -		SFAM01.1	PPB	12/12/2023 18:47	SMPL
O5680-14	E0XS7	509	0.095 -		SFAM01.1	PPB	12/12/2023 18:49	SMPL
O5680-15	E0XS8	2842	0.5071 -		SFAM01.1	PPB	12/12/2023 18:52	SMPL
O5680-16	E0XK9	338	0.0648 -		SFAM01.1	PPB	12/12/2023 18:54	SMPL
O5680-17	E0XG7	269	0.0526 -		SFAM01.1	PPB	12/12/2023 18:56	SMPL
O5680-18	E0XG9	215	0.0431 -		SFAM01.1	PPB	12/12/2023 18:59	SMPL
O5680-19	E0XH0	904	0.1648 -		SFAM01.1	PPB	12/12/2023 19:01	SMPL
CCV092	CCV092	28312	5.0063 -		SFAM01.1	PPB	12/12/2023 19:03	SMPL
CCB092	CCB092	-98	-0.0122 -		SFAM01.1	PPB	12/12/2023 19:05	SMPL
O5680-20	E0XL8	652	0.1202 -		SFAM01.1	PPB	12/12/2023 19:08	SMPL
O5680-21	E0XL8D	609	0.1127 -		SFAM01.1	PPB	12/12/2023 19:10	SMPL
O5680-22	E0XL8S	15636	2.7671 -		SFAM01.1	PPB	12/12/2023 19:12	SMPL
PB157744BL	PBS744	9	0.0067 -		SFAM01.1	PPB	12/12/2023 19:14	SMPL
O4696-03	MDL-SOIL	265	0.0519 -		SFAM01.1	PPB	12/12/2023 19:19	SMPL
CCV093	CCV093	27908	4.935 -		SFAM01.1	PPB	12/12/2023 19:22	SMPL
CCB093	CCB093	-117	-0.0156 -		SFAM01.1	PPB	12/12/2023 19:24	SMPL