



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Sohil

Run Number: LB128704

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/13/2023	10:10
2	CAL2	1	Water	NA	NA	20.3	7.00	12/13/2023	10:11
3	CAL3	1	Water	NA	NA	20.5	10.02	12/13/2023	10:12
4	ICV	1	Water	NA	NA	20.4	7.01	12/13/2023	10:15
5	CCV1	1	Water	NA	NA	20.3	2.01	12/13/2023	10:20
6	O5789-01	1	Solid	20.02	20	21.7	10.80	12/13/2023	10:25
7	O5789-01DUP	1	Solid	20.03	20	21.8	10.81	12/13/2023	10:26
8	O5791-04	1	Solid	20.04	20	20.3	9.39	12/13/2023	10:35
9	O5791-09	1	Solid	20.02	20	20.4	9.83	12/13/2023	10:40
10	CCV2	1	Water	NA	NA	20.6	12.02	12/13/2023	10:45