


LB128708

SFAM01.1

INSTRUMENT ID: CV1

μ Abs.:

52503



Linear

A= 0.0000e+000

B= 1.9042e-004 slope

C= -2.2245e-002 $y = 1.9042x - 2.2245$

Rho= 0.9999380

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.035	-0.035	-65	0.000	-65					
0.05	0.050					442	0				
0.20	0.200	0.193	-0.007	1131	0.0 %	1131					
2.50	2.500	2.509	0.009	13291	0.0 %	13291					
5.00	5.000	5.086	0.086	26827	0.0 %	26827					
7.50	7.500	7.472	-0.028	39356	0.0 %	39356					
10.0	10.000	9.975	-0.025	52503	0.0 %	52503					

LB128708

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0	-65	-		0 SFAM01.1	PPB	12/13/2023 12:28	Std
	0.2	1131	-		0.2 SFAM01.1	PPB	12/13/2023 12:30	Std
	2.5	13291	-		2.5 SFAM01.1	PPB	12/13/2023 12:32	Std
	5	26827	-		5 SFAM01.1	PPB	12/13/2023 12:35	Std
	7.5	39356	-		7.5 SFAM01.1	PPB	12/13/2023 12:37	Std
	10	52503	-		10 SFAM01.1	PPB	12/13/2023 12:39	Std
ICV021	ICV021	20610	3.9022 -		SFAM01.1	PPB	12/13/2023 12:42	SMPL
ICB021	ICB021	-338	-0.0866 -		SFAM01.1	PPB	12/13/2023 12:44	SMPL
CCV098	CCV098	26905	5.1009 -		SFAM01.1	PPB	12/13/2023 12:46	SMPL
CCB098	CCB098	-345	-0.0879 -		SFAM01.1	PPB	12/13/2023 12:49	SMPL
PB157756BL	PBW756	-91	-0.0396 -		SFAM01.1	PPB	12/13/2023 12:51	SMPL
O4696-02	MDL-WATER	344	0.0433 -		SFAM01.1	PPB	12/13/2023 13:09	SMPL
CCV099	CCV099	27107	5.1394 -		SFAM01.1	PPB	12/13/2023 13:11	SMPL
CCB099	CCB099	-275	-0.0746 -		SFAM01.1	PPB	12/13/2023 13:13	SMPL