

LB128715


SFAM01.1

INSTRUMENT ID: CV2

Linear

μ Abs.:

56751



A= 0.0000e+000

B= 1.7780e-004 slope

C= 1.2155e-002 $y = mx - c$

Rho= 0.9998168

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.022	0.022	58	0.000	58					=
0.05	0.050					442	0				16
0.20	0.200	0.231	0.031	1232	0.0 %	1232					-1
2.50	2.500	2.480	-0.020	13880	0.0 %	13880					0
5.00	5.000	4.996	-0.004	28030	0.0 %	28030					-2
7.50	7.500	7.368	-0.132	41369	0.0 %	41369					1
10.0	10.000	10.103	0.103	56751	0.0 %	56751					

LB128715
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0		58	-		0 SFAM01.1	PPB	12/13/2023 16:06	Std
0.2		1232	-		0.2 SFAM01.1	PPB	12/13/2023 16:09	Std
2.5		13880	-		2.5 SFAM01.1	PPB	12/13/2023 16:11	Std
5		28030	-		5 SFAM01.1	PPB	12/13/2023 16:13	Std
7.5		41369	-		7.5 SFAM01.1	PPB	12/13/2023 16:15	Std
10		56751	-		10 SFAM01.1	PPB	12/13/2023 16:18	Std
ICV022	ICV022	21592	3.8513 -		SFAM01.1	PPB	12/13/2023 16:21	SMPL
ICB022	ICB022	-171	-0.0182 -		SFAM01.1	PPB	12/13/2023 16:23	SMPL
CCV001	CCV001	27816	4.9579 -		SFAM01.1	PPB	12/13/2023 16:25	SMPL
CCB001	CCB001	-204	-0.0241 -		SFAM01.1	PPB	12/13/2023 16:28	SMPL
PB157764BL	PBS764	49	0.0209 -		SFAM01.1	PPB	12/13/2023 16:30	SMPL
O4696-04	MDL-SOIL	254	0.0573 -		SFAM01.1	PPB	12/13/2023 16:32	SMPL
CCV002	CCV002	28126	5.013 -		SFAM01.1	PPB	12/13/2023 16:34	SMPL
CCB002	CCB002	-172	-0.0184 -		SFAM01.1	PPB	12/13/2023 16:37	SMPL