



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128722

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	12/14/2023	08:10
2	CAL2	1	Water	NA	NA	20.6	7.01	12/14/2023	08:11
3	CAL3	1	Water	NA	NA	20.5	10.03	12/14/2023	08:15
4	ICV	1	Water	NA	NA	20.5	7.00	12/14/2023	08:19
5	CCV1	1	Water	NA	NA	20.5	2.01	12/14/2023	08:20
6	O5759-03	1	Solid	20.02	20	23.1	8.73	12/14/2023	08:25
7	O5759-03DUP	1	Solid	20.03	20	23.2	8.74	12/14/2023	08:26
8	CCV2	1	Water	NA	NA	20.6	12.02	12/14/2023	08:30

WORKLIST(Hardcopy Internal Chain)

ph 128822

WorkList Name : ph 05759

WorkList ID : 176339

Department : Wet-Chemistry

Date : 12-14-2023 07:53:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O5759-03	AU-06-121323	Solid	pH	Cool 4 deg C	PSEG05	L21	12/13/2023	9045D

Date/Time 12.14.23 08:00
 Raw Sample Received by: ph (cc)
 Raw Sample Relinquished by: JDCSM

Date/Time 12.14.23 11:00
 Raw Sample Received by: JDCSM
 Raw Sample Relinquished by: ph (cc)