



PERCENT SOLID

Supervisor: Iwona  
Analyst: JIGNESH  
Date: 12/15/2023

OVENTEMP IN Celsius(°C): 107  
Time IN: 13:50  
In Date: 12/14/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 07:48  
Out Date: 12/15/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB128725

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|-----------------|--------|----------------|--------------|-------------------------|---------------------------|---------|----------|
| O5807-01 | C0AX3           | 1      | 1.18           | 8.78         | 9.96                    | 4.55                      | 38.4    |          |
| O5807-02 | C0AX4           | 2      | 1.19           | 8.72         | 9.91                    | 2.44                      | 14.3    |          |
| O5807-03 | C0AX5           | 3      | 1.17           | 8.81         | 9.98                    | 5.41                      | 48.1    |          |
| O5807-04 | C0AX6           | 4      | 1.15           | 8.80         | 9.95                    | 4.28                      | 35.6    |          |
| O5807-05 | C0AX7           | 5      | 1.15           | 8.71         | 9.86                    | 7.07                      | 68.0    |          |
| O5807-06 | C0AX8           | 6      | 1.15           | 8.84         | 9.99                    | 7.48                      | 71.6    |          |
| O5807-07 | C0AX9           | 7      | 1.16           | 8.82         | 9.98                    | 7.39                      | 70.6    |          |
| O5807-08 | C0AY0           | 8      | 1.17           | 8.72         | 9.89                    | 5.1                       | 45.1    |          |
| O5807-09 | C0AZ1           | 9      | 1.18           | 8.79         | 9.97                    | 4.87                      | 42.0    |          |
| O5807-10 | C0AZ2           | 10     | 1.18           | 8.37         | 9.55                    | 6.88                      | 68.1    |          |
| O5807-11 | C0AZ3           | 11     | 1.18           | 8.36         | 9.54                    | 4.9                       | 44.5    |          |
| O5807-12 | C0AZ4           | 12     | 1.16           | 8.36         | 9.52                    | 5.74                      | 54.8    |          |
| O5807-13 | C0AZ5           | 13     | 1.17           | 8.46         | 9.63                    | 2.99                      | 21.5    |          |
| O5807-14 | C0AZ6           | 14     | 1.16           | 8.61         | 9.77                    | 8.67                      | 87.2    |          |
| O5807-15 | C0B00           | 15     | 1.16           | 8.80         | 9.96                    | 2.19                      | 11.7    |          |
| O5807-16 | C0B01           | 16     | 1.18           | 8.76         | 9.94                    | 8.61                      | 84.8    |          |
| O5807-17 | C0B18           | 17     | 1.17           | 8.65         | 9.82                    | 6.81                      | 65.2    |          |
| O5807-18 | C0B19           | 18     | 1.17           | 8.65         | 9.82                    | 7.00                      | 67.4    |          |
| O5807-19 | C0B20           | 19     | 1.19           | 8.58         | 9.77                    | 5.66                      | 52.1    |          |
| O5807-20 | VHBLK001        | 20     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | VHBLK    |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$